




# SAR REFERENCE DIPOLE CALIBRATION REPORT

Ref: ACR.353.4.20.M.VGB.A

| Scan Volume | Expanded Uncertainty |
|-------------|----------------------|
| 1 g         | 19 % (SAR)           |
| 10 g        | 19 % (SAR)           |


## 6 CALIBRATION MEASUREMENT RESULTS

### 6.1 RETURN LOSS AND IMPEDANCE IN HEAD LIQUID



| Frequency (MHz) | Return Loss (dB) | Requirement (dB) | Impedance                      |
|-----------------|------------------|------------------|--------------------------------|
| 3500            | -30.21           | -20              | 52.5 $\Omega$ - 1.8 j $\Omega$ |

### 6.2 RETURN LOSS AND IMPEDANCE IN BODY LIQUID



| Frequency (MHz) | Return Loss (dB) | Requirement (dB) | Impedance                      |
|-----------------|------------------|------------------|--------------------------------|
| 3500            | -24.19           | -20              | 56.1 $\Omega$ - 0.3 j $\Omega$ |

Page: 6/13

Template\_ACR.DDD.N.YY.MVGB.ISSUE\_SAR Reference Dipole vG

This document shall not be reproduced, except in full or in part, without the written approval of MVG. The information contained herein is to be used only for the purpose for which it is submitted and is not to be released in whole or part without written approval of MVG.



# SAR REFERENCE DIPOLE CALIBRATION REPORT

Ref: ACR.353.4.20.M VGB.A

## 6.3 MECHANICAL DIMENSIONS

| Frequency MHz | L mm      |          | h mm      |          | d mm     |          |
|---------------|-----------|----------|-----------|----------|----------|----------|
|               | required  | measured | required  | measured | required | measured |
| 300           | 420.0±1 % |          | 250.0±1 % |          | 6.35±1 % |          |
| 450           | 290.0±1 % |          | 166.7±1 % |          | 6.35±1 % |          |
| 750           | 176.0±1 % |          | 100.0±1 % |          | 6.35±1 % |          |
| 835           | 161.0±1 % |          | 89.8±1 %  |          | 3.6±1 %  |          |
| 900           | 149.0±1 % |          | 83.3±1 %  |          | 3.6±1 %  |          |
| 1450          | 89.1±1 %  |          | 51.7±1 %  |          | 3.6±1 %  |          |
| 1500          | 80.5±1 %  |          | 50.0±1 %  |          | 3.6±1 %  |          |
| 1640          | 79.0±1 %  |          | 45.7±1 %  |          | 3.6±1 %  |          |
| 1750          | 75.2±1 %  |          | 42.9±1 %  |          | 3.6±1 %  |          |
| 1800          | 72.0±1 %  |          | 41.7±1 %  |          | 3.6±1 %  |          |
| 1900          | 68.0±1 %  |          | 39.5±1 %  |          | 3.6±1 %  |          |
| 1950          | 66.3±1 %  |          | 38.5±1 %  |          | 3.6±1 %  |          |
| 2000          | 64.5±1 %  |          | 37.5±1 %  |          | 3.6±1 %  |          |
| 2100          | 61.0±1 %  |          | 35.7±1 %  |          | 3.6±1 %  |          |
| 2300          | 55.5±1 %  |          | 32.6±1 %  |          | 3.6±1 %  |          |
| 2450          | 51.5±1 %  |          | 30.4±1 %  |          | 3.6±1 %  |          |
| 2600          | 48.5±1 %  |          | 28.8±1 %  |          | 3.6±1 %  |          |
| 3000          | 41.5±1 %  |          | 25.0±1 %  |          | 3.6±1 %  |          |
| 3300          | -         |          | -         |          | -        |          |
| 3500          | 37.0±1 %  | 37.20    | 26.4±1 %  | 26.34    | 3.6±1 %  | 3.60     |
| 3700          | 34.7±1 %  |          | 26.4±1 %  |          | 3.6±1 %  |          |
| 3900          | -         |          | -         |          | -        |          |
| 4200          | -         |          | -         |          | -        |          |
| 4600          | -         |          | -         |          | -        |          |
| 4900          | -         |          | -         |          | -        |          |

## 7 VALIDATION MEASUREMENT

The IEEE Std. 1528, FCC KDBs and CEI/IEC 62209 standards state that the system validation measurements must be performed using a reference dipole meeting the fore mentioned return loss and mechanical dimension requirements. The validation measurement must be performed against a liquid filled flat phantom, with the phantom constructed as outlined in the fore mentioned standards. Per the standards, the dipole shall be positioned below the bottom of the phantom, with the dipole length centered and parallel to the longest dimension of the flat phantom, with the top surface of the dipole at the described distance from the bottom surface of the phantom.

Page: 7/13

Template\_ACR.DDD.N.YY.MVGB.ISSUE\_SAR Reference Dipole vG

This document shall not be reproduced, except in full or in part, without the written approval of MVG. The information contained herein is to be used only for the purpose for which it is submitted and is not to be released in whole or part without written approval of MVG.



## SAR REFERENCE DIPOLE CALIBRATION REPORT

Ref: ACR.353.4.20.M VGB.A

### 7.1 HEAD LIQUID MEASUREMENT

| Frequency<br>MHz | Relative permittivity ( $\epsilon_r$ ) |          | Conductivity ( $\sigma$ ) S/m |          |
|------------------|----------------------------------------|----------|-------------------------------|----------|
|                  | required                               | measured | required                      | measured |
| 300              | 45.3 $\pm$ 10 %                        |          | 0.87 $\pm$ 10 %               |          |
| 450              | 43.5 $\pm$ 10 %                        |          | 0.87 $\pm$ 10 %               |          |
| 750              | 41.9 $\pm$ 10 %                        |          | 0.89 $\pm$ 10 %               |          |
| 835              | 41.5 $\pm$ 10 %                        |          | 0.90 $\pm$ 10 %               |          |
| 900              | 41.5 $\pm$ 10 %                        |          | 0.97 $\pm$ 10 %               |          |
| 1450             | 40.5 $\pm$ 10 %                        |          | 1.20 $\pm$ 10 %               |          |
| 1500             | 40.4 $\pm$ 10 %                        |          | 1.23 $\pm$ 10 %               |          |
| 1640             | 40.2 $\pm$ 10 %                        |          | 1.31 $\pm$ 10 %               |          |
| 1750             | 40.1 $\pm$ 10 %                        |          | 1.37 $\pm$ 10 %               |          |
| 1800             | 40.0 $\pm$ 10 %                        |          | 1.40 $\pm$ 10 %               |          |
| 1900             | 40.0 $\pm$ 10 %                        |          | 1.40 $\pm$ 10 %               |          |
| 1950             | 40.0 $\pm$ 10 %                        |          | 1.40 $\pm$ 10 %               |          |
| 2000             | 40.0 $\pm$ 10 %                        |          | 1.40 $\pm$ 10 %               |          |
| 2100             | 39.8 $\pm$ 10 %                        |          | 1.49 $\pm$ 10 %               |          |
| 2300             | 39.5 $\pm$ 10 %                        |          | 1.67 $\pm$ 10 %               |          |
| 2450             | 39.2 $\pm$ 10 %                        |          | 1.80 $\pm$ 10 %               |          |
| 2600             | 39.0 $\pm$ 10 %                        |          | 1.96 $\pm$ 10 %               |          |
| 3000             | 38.5 $\pm$ 10 %                        |          | 2.40 $\pm$ 10 %               |          |
| 3300             | 38.2 $\pm$ 10 %                        |          | 2.71 $\pm$ 10 %               |          |
| 3500             | 37.9 $\pm$ 10 %                        | 36.5     | 2.91 $\pm$ 10 %               | 3.07     |
| 3700             | 37.7 $\pm$ 10 %                        |          | 3.12 $\pm$ 10 %               |          |
| 3900             | 37.5 $\pm$ 10 %                        |          | 3.32 $\pm$ 10 %               |          |
| 4200             | 37.1 $\pm$ 10 %                        |          | 3.63 $\pm$ 10 %               |          |
| 4600             | 36.7 $\pm$ 10 %                        |          | 4.04 $\pm$ 10 %               |          |
| 4900             | 36.3 $\pm$ 10 %                        |          | 4.35 $\pm$ 10 %               |          |

### 7.2 SAR MEASUREMENT RESULT WITH HEAD LIQUID

The IEEE Std. 1528 and CEI/IEC 62209 standards state that the system validation measurements should produce the SAR values shown below (for phantom thickness of 2 mm), within the uncertainty for the system validation. All SAR values are normalized to 1 W forward power. In bracket, the measured SAR is given with the used input power.

Page: 8/13

Template\_ACR.DDD.N.YY.MVGB.ISSUE\_SAR Reference Dipole vG

This document shall not be reproduced, except in full or in part, without the written approval of MVG. The information contained herein is to be used only for the purpose for which it is submitted and is not to be released in whole or part without written approval of MVG.



## SAR REFERENCE DIPOLE CALIBRATION REPORT

Ref: ACR.353.4.20.M.VGB.A

|                                           |                                                          |
|-------------------------------------------|----------------------------------------------------------|
| Software                                  | OPENSAR V5                                               |
| Phantom                                   | SN 13/09 SAM68                                           |
| Probe                                     | SN 41/18 EPG0333                                         |
| Liquid                                    | Head Liquid Values: $\epsilon_{ps}'$ : 36.5 sigma : 3.07 |
| Distance between dipole center and liquid | 10.0 mm                                                  |
| Area scan resolution                      | $dx=8mm/dy=8mm$                                          |
| Zoon Scan Resolution                      | $dx=5mm/dy=5mm/dz=4mm$                                   |
| Frequency                                 | 3500 MHz                                                 |
| Input power                               | 20 dBm                                                   |
| Liquid Temperature                        | 20 +/- 1 °C                                              |
| Lab Temperature                           | 20 +/- 1 °C                                              |
| Lab Humidity                              | 30-70 %                                                  |

| Frequency<br>MHz | 1 g SAR (W/kg/W) |              | 10 g SAR (W/kg/W) |              |
|------------------|------------------|--------------|-------------------|--------------|
|                  | required         | measured     | required          | measured     |
| 300              | 2.85             |              | 1.94              |              |
| 450              | 4.58             |              | 3.06              |              |
| 750              | 8.49             |              | 5.55              |              |
| 835              | 9.56             |              | 6.22              |              |
| 900              | 10.9             |              | 6.99              |              |
| 1450             | 29               |              | 16                |              |
| 1500             | 30.5             |              | 16.8              |              |
| 1640             | 34.2             |              | 18.4              |              |
| 1750             | 36.4             |              | 19.3              |              |
| 1800             | 38.4             |              | 20.1              |              |
| 1900             | 39.7             |              | 20.5              |              |
| 1950             | 40.5             |              | 20.9              |              |
| 2000             | 41.1             |              | 21.1              |              |
| 2100             | 43.6             |              | 21.9              |              |
| 2300             | 48.7             |              | 23.3              |              |
| 2450             | 52.4             |              | 24                |              |
| 2600             | 55.3             |              | 24.6              |              |
| 3000             | 63.8             |              | 25.7              |              |
| 3300             | -                |              | -                 |              |
| 3500             | 67.1             | 67.28 (6.73) | 25                | 24.80 (2.48) |
| 3700             | 67.4             |              | 24.2              |              |
| 3900             | -                |              | -                 |              |
| 4200             | -                |              | -                 |              |
| 4600             | -                |              | -                 |              |
| 4900             | -                |              | -                 |              |

Page: 5/13


Template\_ACR.DDD.N.YY.MVGB.ISSUE\_SAR Reference Dipole vG

This document shall not be reproduced, except in full or in part, without the written approval of MVG. The information contained herein is to be used only for the purpose for which it is submitted and is not to be released in whole or part without written approval of MVG.



# SAR REFERENCE DIPOLE CALIBRATION REPORT

Ref: ACR.353.4.20.MVGB.A



Page: 16/13

**Template\_ACR.DDD.N.YY.MVGB.ISSUE\_SAR Reference Dipole vG**

*This document shall not be reproduced, except in full or in part, without the written approval of MVG. The information contained herein is to be used only for the purpose for which it is submitted and is not to be released in whole or part without written approval of MVG.*





## SAR REFERENCE DIPOLE CALIBRATION REPORT

Ref: ACR.353.4.20.M.VGB.A

## 7.3 BODY LIQUID MEASUREMENT

| Frequency<br>MHz | Relative permittivity ( $\epsilon_r$ ) |          | Conductivity ( $\sigma$ ) S/m |          |
|------------------|----------------------------------------|----------|-------------------------------|----------|
|                  | required                               | measured | required                      | measured |
| 150              | 61.9 $\pm$ 10 %                        |          | 0.80 $\pm$ 10 %               |          |
| 300              | 58.2 $\pm$ 10 %                        |          | 0.92 $\pm$ 10 %               |          |
| 450              | 56.7 $\pm$ 10 %                        |          | 0.94 $\pm$ 10 %               |          |
| 750              | 55.5 $\pm$ 10 %                        |          | 0.96 $\pm$ 10 %               |          |
| 835              | 55.2 $\pm$ 10 %                        |          | 0.97 $\pm$ 10 %               |          |
| 900              | 55.0 $\pm$ 10 %                        |          | 1.05 $\pm$ 10 %               |          |
| 915              | 55.0 $\pm$ 10 %                        |          | 1.06 $\pm$ 10 %               |          |
| 1450             | 54.0 $\pm$ 10 %                        |          | 1.30 $\pm$ 10 %               |          |
| 1610             | 53.8 $\pm$ 10 %                        |          | 1.40 $\pm$ 10 %               |          |
| 1800             | 53.3 $\pm$ 10 %                        |          | 1.52 $\pm$ 10 %               |          |
| 1900             | 53.3 $\pm$ 10 %                        |          | 1.52 $\pm$ 10 %               |          |
| 2000             | 53.3 $\pm$ 10 %                        |          | 1.52 $\pm$ 10 %               |          |
| 2100             | 53.2 $\pm$ 10 %                        |          | 1.62 $\pm$ 10 %               |          |
| 2300             | 52.9 $\pm$ 10 %                        |          | 1.81 $\pm$ 10 %               |          |
| 2450             | 52.7 $\pm$ 10 %                        |          | 1.95 $\pm$ 10 %               |          |
| 2600             | 52.5 $\pm$ 10 %                        |          | 2.16 $\pm$ 10 %               |          |
| 3000             | 52.0 $\pm$ 10 %                        |          | 2.73 $\pm$ 10 %               |          |
| 3300             | 51.6 $\pm$ 10 %                        |          | 3.08 $\pm$ 10 %               |          |
| 3500             | 51.3 $\pm$ 10 %                        | 48.6     | 3.31 $\pm$ 10 %               | 3.29     |
| 3700             | 51.0 $\pm$ 10 %                        |          | 3.55 $\pm$ 10 %               |          |
| 3900             | 50.8 $\pm$ 10 %                        |          | 3.78 $\pm$ 10 %               |          |
| 4200             | 50.4 $\pm$ 10 %                        |          | 4.13 $\pm$ 10 %               |          |
| 4600             | 49.8 $\pm$ 10 %                        |          | 4.60 $\pm$ 10 %               |          |
| 4900             | 49.4 $\pm$ 10 %                        |          | 4.95 $\pm$ 10 %               |          |
| 5200             | 49.0 $\pm$ 10 %                        |          | 5.30 $\pm$ 10 %               |          |
| 5300             | 48.9 $\pm$ 10 %                        |          | 5.42 $\pm$ 10 %               |          |
| 5400             | 48.7 $\pm$ 10 %                        |          | 5.53 $\pm$ 10 %               |          |
| 5500             | 48.6 $\pm$ 10 %                        |          | 5.65 $\pm$ 10 %               |          |
| 5600             | 48.5 $\pm$ 10 %                        |          | 5.77 $\pm$ 10 %               |          |
| 5800             | 48.2 $\pm$ 10 %                        |          | 6.00 $\pm$ 10 %               |          |

Page: 11/13

Template\_ACR.DDD.N.YY.MVGB.ISSUE\_SAR Reference Dipole vG

This document shall not be reproduced, except in full or in part, without the written approval of MVG. The information contained herein is to be used only for the purpose for which it is submitted and is not to be released in whole or part without written approval of MVG.




# SAR REFERENCE DIPOLE CALIBRATION REPORT

Ref: ACR.353.4.20.M.VGB.A

## 7.4 SAR MEASUREMENT RESULT WITH BODY LIQUID

|                                           |                                                           |
|-------------------------------------------|-----------------------------------------------------------|
| Software                                  | OPENSAR V5                                                |
| Phantom                                   | SN 13/09 SAM68                                            |
| Probe                                     | SN 41/18 EPG0333                                          |
| Liquid                                    | Body Liquid Values: $\epsilon_{ps}^*$ : 48.6 sigma : 3.29 |
| Distance between dipole center and liquid | 10.0 mm                                                   |
| Area scan resolution                      | $dx=8mm/dy=8mm$                                           |
| Zoom Scan Resolution                      | $dx=5mm/dy=5mm/dz=4mm$                                    |
| Frequency                                 | 3500 MHz                                                  |
| Input power                               | 20 dBm                                                    |
| Liquid Temperature                        | 20 +/- 1 °C                                               |
| Lab Temperature                           | 20 +/- 1 °C                                               |
| Lab Humidity                              | 30-70 %                                                   |

| Frequency MHz | 1 g SAR (W/kg/W) | 10 g SAR (W/kg/W) |
|---------------|------------------|-------------------|
|               | measured         | measured          |
| 3500          | 64.70 (6.47)     | 24.04 (2.40)      |





## SAR REFERENCE DIPOLE CALIBRATION REPORT

Ref: ACR.353.4.20.M.VGB.A

## 8 LIST OF EQUIPMENT

| Equipment Summary Sheet            |                         |                    |                                               |                                               |
|------------------------------------|-------------------------|--------------------|-----------------------------------------------|-----------------------------------------------|
| Equipment Description              | Manufacturer / Model    | Identification No. | Current Calibration Date                      | Next Calibration Date                         |
| SAM Phantom                        | MVG                     | SN-13/09-SAM6B     | Validated. No cal required.                   | Validated. No cal required.                   |
| COMOSAR Test Bench                 | Version 3               | NA                 | Validated. No cal required.                   | Validated. No cal required.                   |
| Network Analyzer                   | Rohde & Schwarz ZVM     | 100203             | 05/2019                                       | 05/2022                                       |
| Network Analyzer – Calibration kit | Rohde & Schwarz ZV-Z235 | 101223             | 05/2019                                       | 05/2022                                       |
| Calipers                           | Mitutoyo                | SN 0009732         | 10/2019                                       | 10/2022                                       |
| Reference Probe                    | MVG                     | EPGO333 SN 41/1B   | 05/2020                                       | 05/2021                                       |
| Multimeter                         | Keithley 2000           | 1160271            | 02/2020                                       | 02/2023                                       |
| Signal Generator                   | Rohde & Schwarz SMB     | 106589             | 04/2019                                       | 04/2022                                       |
| Amplifier                          | Aethercomm              | SN 046             | Characterized prior to test. No cal required. | Characterized prior to test. No cal required. |
| Power Meter                        | NI-USB 5680             | 170100013          | 05/2019                                       | 05/2022                                       |
| Directional Coupler                | Narda 4216-20           | 01386              | Characterized prior to test. No cal required. | Characterized prior to test. No cal required. |
| Temperature / Humidity Sensor      | Testo 184 H1            | 44220687           | 05/2020                                       | 05/2023                                       |

Page: 13/13

Template\_ACR.DDD.N.YY.MVGB.ISSUE\_SAR Reference Dipole vG

This document shall not be reproduced, except in full or in part, without the written approval of MVG. The information contained herein is to be used only for the purpose for which it is submitted and is not to be released in whole or part without written approval of MVG.



## SID5G Dipole Calibration Report

**SAR Reference Waveguide Calibration Report**

Ref : ACR.178.10.20.MVGB.A

**CCIC SOUTHERN TESTING CO., LTD**  
**ELECTRONIC TESTING BUILDING, NO. 43 SHAHE**  
**ROAD, XILI STREET, NANSHAN**  
**DISTRICT SHENZHEN, GUANGDONG, CHINA**  
**MVG**  
**COMOSAR REFERENCE WAVEGUIDE**  
**FREQUENCY: 5000-6000 MHZ**  
**SERIAL NO.: SN 15/15 WGA39**

**Calibrated at MVG**  
**Z.I. de la pointe du diable**  
**Technopôle Brest Iroise – 295 avenue Alexis de Rochon**  
**29280 PLOUZANE - FRANCE**

**Calibration date: 06/25/2020**

Accreditations #2-6789 and #2-6814  
Seal available on [www.cofrac.fr](http://www.cofrac.fr)


**Summary:**

This document presents the method and results from an accredited SAR reference waveguide calibration performed at MVG, using the COMOSAR test bench. The test results covered by accreditation are traceable to the International System of Units (SI).



## SAR REFERENCE WAVEGUIDE CALIBRATION REPORT

Ref: ACR.178.10.20.MVGB-A

|               | Name         | Function            | Date      | Signature                                                                           |
|---------------|--------------|---------------------|-----------|-------------------------------------------------------------------------------------|
| Prepared by : | Jérôme LUC   | Technical Manager   | 6/26/2020 |  |
| Checked by :  | Jérôme LUC   | Technical Manager   | 6/26/2020 |  |
| Approved by : | Yann Toutain | Laboratory Director | 6/26/2020 |  |

|                | Customer Name                        |
|----------------|--------------------------------------|
| Distribution : | CCIC SOUTHERN<br>TESTING CO.,<br>LTD |

| Issue | Name           | Date      | Modifications   |
|-------|----------------|-----------|-----------------|
| A     | Jérôme LE GALL | 6/26/2020 | Initial release |
|       |                |           |                 |
|       |                |           |                 |
|       |                |           |                 |

Page: 2/13

Template: ACR.DDD.N.YY.MVGR.ISSUE\_SAR Reference Waveguide vG

This document shall not be reproduced, except in full or in part, without the written approval of MVG. The information contained herein is to be used only for the purpose for which it is submitted and is not to be released in whole or part without written approval of MVG.



## TABLE OF CONTENTS

|     |                                      |    |
|-----|--------------------------------------|----|
| 1   | Introduction.....                    | 4  |
| 2   | Device Under Test .....              | 4  |
| 3   | Product Description .....            | 4  |
| 3.1 | General Information .....            | 4  |
| 4   | Measurement Method .....             | 4  |
| 4.1 | Return Loss Requirements .....       | 4  |
| 4.2 | Mechanical Requirements .....        | 4  |
| 5   | Measurement Uncertainty .....        | 4  |
| 5.1 | Return Loss .....                    | 5  |
| 5.2 | Dimension Measurement .....          | 5  |
| 5.3 | Validation Measurement .....         | 5  |
| 6   | Calibration Measurement Results..... | 5  |
| 6.1 | Return Loss .....                    | 5  |
| 6.2 | Mechanical Dimensions .....          | 6  |
| 7   | Validation measurement .....         | 7  |
| 7.1 | Head Liquid Measurement .....        | 7  |
| 7.2 | Measurement Result .....             | 7  |
| 7.3 | Body Measurement Result .....        | 10 |
| 8   | List of Equipment .....              | 13 |



## 1 INTRODUCTION

This document contains a summary of the requirements set forth by the IEEE 1528 and CEI/IEC 62209 standards for reference waveguides used for SAR measurement system validations and the measurements that were performed to verify that the product complies with the fore mentioned standards.

## 2 DEVICE UNDER TEST

| Device Under Test              |                                           |
|--------------------------------|-------------------------------------------|
| Device Type                    | COMOSAR 5000-6000 MHz REFERENCE WAVEGUIDE |
| Manufacturer                   | MVG                                       |
| Model                          | SWG5500                                   |
| Serial Number                  | SN 15/15 WGA39                            |
| Product Condition (new / used) | Used                                      |

## 3 PRODUCT DESCRIPTION

### 3.1 GENERAL INFORMATION

MVG's COMOSAR Validation Waveguides are built in accordance to the IEEE 1528 and CEI/IEC 62209 standards.

## 4 MEASUREMENT METHOD

The IEEE 1528 and CEI/IEC 62209 standards provide requirements for reference waveguides used for system validation measurements. The following measurements were performed to verify that the product complies with the fore mentioned standards.

### 4.1 RETURN LOSS REQUIREMENTS

The waveguide used for SAR system validation measurements and checks must have a return loss of -8 dB or better. The return loss measurement shall be performed with matching layer placed in the open end of the waveguide, with the waveguide and matching layer in direct contact with the phantom shell as outlined in the fore mentioned standards. A direct method is used with a network analyser and its calibration kit, both with a valid ISO17025 calibration.

### 4.2 MECHANICAL REQUIREMENTS

The IEEE 1528 and CEI/IEC 62209 standards specify the mechanical dimensions of the validation waveguide, the specified dimensions are as shown in Section 6.2. Figure 1 shows how the dimensions relate to the physical construction of the waveguide. A direct method is used with a ISO17025 calibrated caliper.

Page: 4/13

*Template: ACR.DDD.N.YY.MVGB.ISSUE\_SAR Reference Waveguide.vG*

*This document shall not be reproduced, except in full or in part, without the written approval of MVG. The information contained herein is to be used only for the purpose for which it is submitted and is not to be released in whole or part without written approval of MVG.*



## 5 MEASUREMENT UNCERTAINTY

All uncertainties listed below represent an expanded uncertainty expressed at approximately the 95% confidence level using a coverage factor of  $k=2$ , traceable to the Internationally Accepted Guides to Measurement Uncertainty.

### 5.1 RETURN LOSS

The following uncertainties apply to the return loss measurement:

| Frequency band | Expanded Uncertainty on Return Loss |
|----------------|-------------------------------------|
| 400-6000MHz    | 0.08 LIN                            |

### 5.2 DIMENSION MEASUREMENT

The following uncertainties apply to the dimension measurements:

| Length (mm) | Expanded Uncertainty on Length |
|-------------|--------------------------------|
| 0 - 300     | 0.20 mm                        |


### 5.3 VALIDATION MEASUREMENT

The guidelines outlined in the IEEE 1528 and CEI/IEC 62209 standards were followed to generate the measurement uncertainty for validation measurements.

| Scan Volume | Expanded Uncertainty |
|-------------|----------------------|
| 1 g         | 19 % (SAR)           |
| 10 g        | 19 % (SAR)           |

## 6 CALIBRATION MEASUREMENT RESULTS

### 6.1 RETURN LOSS IN HEAD LIQUID



Page: 5/13

Template: ACR.DDD.N.YY.MVGB.ISSUE\_SAR Reference Waveguide vG

This document shall not be reproduced, except in full or in part, without the written approval of MVG. The information contained herein is to be used only for the purpose for which it is submitted and is not to be released in whole or part without written approval of MVG.






# SAR REFERENCE WAVEGUIDE CALIBRATION REPORT

Ref: ACR.178.10.20.MVGB.A

| Frequency (MHz) | Return Loss (dB) | Requirement (dB) | Impedance                      |
|-----------------|------------------|------------------|--------------------------------|
| 5200            | -9.80            | -8               | $22.19 \Omega + 10.27 j\Omega$ |
| 5400            | -11.30           | -8               | $75.26 \Omega + 2.27 j\Omega$  |
| 5600            | -14.62           | -8               | $34.93 \Omega - 10.02 j\Omega$ |
| 5800            | -13.85           | -8               | $55.11 \Omega + 19.16 j\Omega$ |

## 6.2 RETURN LOSS IN BODY LIQUID



| Frequency (MHz) | Return Loss (dB) | Requirement (dB) | Impedance                      |
|-----------------|------------------|------------------|--------------------------------|
| 5200            | -9.32            | -8               | $20.88 \Omega + 10.83 j\Omega$ |
| 5400            | -10.98           | -8               | $76.16 \Omega - 0.57 j\Omega$  |
| 5600            | -13.52           | -8               | $31.86 \Omega - 9.17 j\Omega$  |
| 5800            | -13.05           | -8               | $57.79 \Omega + 20.13 j\Omega$ |

## 6.3 MECHANICAL DIMENSIONS

| Frequency (MHz) | L (mm)       |          | W (mm)       |          | L <sub>r</sub> (mm) |          | W <sub>r</sub> (mm) |          |
|-----------------|--------------|----------|--------------|----------|---------------------|----------|---------------------|----------|
|                 | Required     | Measured | Required     | Measured | Required            | Measured | Required            | Measured |
| 5800            | 40.39 ± 0.13 | -        | 20.19 ± 0.13 | -        | 81.03 ± 0.13        | -        | 61.98 ± 0.13        | -        |

Page: 6/13

Template: ACR.DDD/N.YY.MVGB.ISSUE\_SAR Reference Waveguide vG

This document shall not be reproduced, except in full or in part, without the written approval of MVG. The information contained herein is to be used only for the purpose for which it is submitted and is not to be released in whole or part without written approval of MVG.




Figure 1: Validation Waveguide Dimensions

## 7 VALIDATION MEASUREMENT

The IEEE Std. 1528 and CEI/IEC 62209 standards state that the system validation measurements must be performed using a reference waveguide meeting the fore mentioned return loss and mechanical dimension requirements. The validation measurement must be performed with the matching layer placed in the open end of the waveguide, with the waveguide and matching layer in direct contact with the phantom shell.

### 7.1 HEAD LIQUID MEASUREMENT

| Frequency<br>MHz | Relative permittivity ( $\epsilon_r'$ ) |          | Conductivity ( $\sigma$ ) S/m |          |
|------------------|-----------------------------------------|----------|-------------------------------|----------|
|                  | required                                | measured | required                      | measured |
| 5000             | 36.2 $\pm$ 10 %                         |          | 4.45 $\pm$ 10 %               |          |
| 5100             | 36.1 $\pm$ 10 %                         |          | 4.56 $\pm$ 10 %               |          |
| 5200             | 36.0 $\pm$ 10 %                         | 34.60    | 4.66 $\pm$ 10 %               | 4.55     |
| 5300             | 35.9 $\pm$ 10 %                         |          | 4.76 $\pm$ 10 %               |          |
| 5400             | 35.8 $\pm$ 10 %                         | 34.02    | 4.86 $\pm$ 10 %               | 4.88     |
| 5500             | 35.6 $\pm$ 10 %                         |          | 4.97 $\pm$ 10 %               |          |
| 5600             | 35.5 $\pm$ 10 %                         | 33.46    | 5.07 $\pm$ 10 %               | 5.25     |
| 5700             | 35.4 $\pm$ 10 %                         |          | 5.17 $\pm$ 10 %               |          |
| 5800             | 35.3 $\pm$ 10 %                         | 32.78    | 5.27 $\pm$ 10 %               | 5.64     |
| 5900             | 35.2 $\pm$ 10 %                         |          | 5.38 $\pm$ 10 %               |          |
| 6000             | 35.1 $\pm$ 10 %                         |          | 5.48 $\pm$ 10 %               |          |

### 7.2 SAR MEASUREMENT RESULT WITH HEAD LIQUID

At those frequencies, the target SAR value can not be generic. Hereunder is the target SAR value defined by MVG, within the uncertainty for the system validation. All SAR values are normalized to 1 W net power. In bracket, the measured SAR is given with the used input power.

Page: 7/13

Template: ACR.DDD.N.YY.MVGB.ISSUE\_SAR Reference Waveguide vG

This document shall not be reproduced, except in full or in part, without the written approval of MVG. The information contained herein is to be used only for the purpose for which it is submitted and is not to be released in whole or part without written approval of MVG.




## SAR REFERENCE WAVEGUIDE CALIBRATION REPORT

Ref: ACR.178.10.20.MVGB.A

|                                              |                                                                                                                                                                                                                              |
|----------------------------------------------|------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
| Software                                     | OPENSAR V5                                                                                                                                                                                                                   |
| Phantom                                      | SN 13/09 SAM68                                                                                                                                                                                                               |
| Probe                                        | SN 41/18 EPG0333                                                                                                                                                                                                             |
| Liquid                                       | Head Liquid Values 5200 MHz: eps':34.60 sigma : 4.55<br>Head Liquid Values 5400 MHz: eps':34.02 sigma : 4.88<br>Head Liquid Values 5600 MHz: eps':33.46 sigma : 5.25<br>Head Liquid Values 5800 MHz: eps':32.78 sigma : 5.64 |
| Distance between dipole waveguide and liquid | 0 mm                                                                                                                                                                                                                         |
| Area scan resolution                         | dx=8mm/dy=8mm                                                                                                                                                                                                                |
| Zoon Scan Resolution                         | dx=4mm/dy=4m/dz=2mm                                                                                                                                                                                                          |
| Frequency                                    | 5200 MHz<br>5400 MHz<br>5600 MHz<br>5800 MHz                                                                                                                                                                                 |
| Input power                                  | 20 dBm                                                                                                                                                                                                                       |
| Liquid Temperature                           | 20 +/- 1 °C                                                                                                                                                                                                                  |
| Lab Temperature                              | 20 +/- 1 °C                                                                                                                                                                                                                  |
| Lab Humidity                                 | 30-70 %                                                                                                                                                                                                                      |

| Frequency (MHz) | 1 g SAR (W/kg) |                | 10 g SAR (W/kg) |              |
|-----------------|----------------|----------------|-----------------|--------------|
|                 | required       | measured       | required        | measured     |
| 5200            | 159.00         | 151.11 (15.11) | 56.90           | 54.03 (5.40) |
| 5400            | 166.40         | 159.92 (15.99) | 58.43           | 56.78 (5.68) |
| 5600            | 173.80         | 165.99 (16.60) | 59.97           | 58.35 (5.83) |
| 5800            | 181.20         | 176.86 (17.69) | 61.50           | 61.84 (6.18) |

## SAR MEASUREMENT PLOTS @ 5200 MHz



Page: 8/13

Template: ACR.DDD.N.V1.MVGB.ISSUE\_SAR Reference Waveguide vG


This document shall not be reproduced, except in full or in part, without the written approval of MVQ. The information contained herein is to be used only for the purpose for which it is submitted and is not to be released in whole or part without written approval of MVQ.




# SAR REFERENCE WAVEGUIDE CALIBRATION REPORT

Ref: ACR.178.10.20.MVGB.A


## SAR MEASUREMENT PLOTS @ 5400 MHz



## SAR MEASUREMENT PLOTS @ 5600 MHz



## SAR MEASUREMENT PLOTS @ 5800 MHz



Page: 9/13

Template: ACR.DDD.N.YY.MVGB.ISSUE\_SAR Reference Waveguide vG

This document shall not be reproduced, except in full or in part, without the written approval of MVG. The information contained herein is to be used only for the purpose for which it is submitted and is not to be released in whole or part without written approval of MVG.





## SAR REFERENCE WAVEGUIDE CALIBRATION REPORT

Ref: ACR.178.10.25.MVGBLA

### 7.3 BODY LIQUID MEASUREMENT

| Frequency<br>MHz | Relative permittivity [ $\epsilon_r$ ] |          | Conductivity [ $\sigma$ ] S/m |          |
|------------------|----------------------------------------|----------|-------------------------------|----------|
|                  | required                               | measured | required                      | measured |
| 5200             | 49.0 $\pm$ 10 %                        | 45.25    | 5.30 $\pm$ 10 %               | 5.42     |
| 5300             | 48.9 $\pm$ 10 %                        |          | 5.42 $\pm$ 10 %               |          |
| 5400             | 48.7 $\pm$ 10 %                        | 45.09    | 5.53 $\pm$ 10 %               | 5.80     |
| 5500             | 48.6 $\pm$ 10 %                        |          | 5.65 $\pm$ 10 %               |          |
| 5600             | 48.5 $\pm$ 10 %                        | 44.84    | 5.77 $\pm$ 10 %               | 6.20     |
| 5800             | 48.2 $\pm$ 10 %                        | 44.59    | 6.00 $\pm$ 10 %               | 6.56     |

### 7.4 SAR MEASUREMENT RESULT WITH BODY LIQUID

|                                              |                                                                                                                                                                                                                                                                  |
|----------------------------------------------|------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
| Software                                     | OPENSAR V5                                                                                                                                                                                                                                                       |
| Phantom                                      | SN 13/09 SAM68                                                                                                                                                                                                                                                   |
| Probe                                        | SN 41/18 EPG0333                                                                                                                                                                                                                                                 |
| Liquid                                       | Body Liquid Values 5200 MHz: $\epsilon_r$ :45.25 sigma : 5.42<br>Body Liquid Values 5400 MHz: $\epsilon_r$ :45.09 sigma : 5.80<br>Body Liquid Values 5600 MHz: $\epsilon_r$ :44.84 sigma : 6.20<br>Body Liquid Values 5800 MHz: $\epsilon_r$ :44.59 sigma : 6.56 |
| Distance between dipole waveguide and liquid | 0 mm                                                                                                                                                                                                                                                             |
| Area scan resolution                         | $\Delta x=8\text{mm}/\Delta y=8\text{mm}$                                                                                                                                                                                                                        |
| Zoon Scan Resolution                         | $\Delta x=4\text{mm}/\Delta y=4\text{mm}/\Delta z=2\text{mm}$                                                                                                                                                                                                    |
| Frequency                                    | 5200 MHz<br>5400 MHz<br>5600 MHz<br>5800 MHz                                                                                                                                                                                                                     |
| Input power                                  | 20 dBm                                                                                                                                                                                                                                                           |
| Liquid Temperature                           | 20 $\pm$ 1 $^{\circ}\text{C}$                                                                                                                                                                                                                                    |
| Lab Temperature                              | 20 $\pm$ 1 $^{\circ}\text{C}$                                                                                                                                                                                                                                    |
| Lab Humidity                                 | 30-70 %                                                                                                                                                                                                                                                          |

| Frequency (MHz) | 1 g SAR (W/kg)<br>measured | 10 g SAR (W/kg)<br>measured |
|-----------------|----------------------------|-----------------------------|
| 5200            | 150.13 (15.01)             | 55.47 (5.55)                |
| 5400            | 152.16 (15.22)             | 55.69 (5.57)                |
| 5600            | 154.92 (15.49)             | 56.02 (5.60)                |
| 5800            | 158.10 (15.81)             | 54.64 (5.46)                |

Page: 10/13

Template: ACR.DDD.N.YY.MVGB.ISSUE SAR Reference Waveguide vG

This document shall not be reproduced, except in full or in part, without the written approval of MVG. The information contained herein is to be used only for the purpose for which it is submitted and is not to be released in whole or part without written approval of MVG.







## SAR REFERENCE WAVEGUIDE CALIBRATION REPORT

Ref: ACR.178.10.20.MVGB.A


### BODY SAR MEASUREMENT PLOTS @ 5200 MHz



### BODY SAR MEASUREMENT PLOTS @ 5400 MHz



### BODY SAR MEASUREMENT PLOTS @ 5600 MHz



Page: 11/13

Template: ACR.DDB\N.Y\MVGB.ISSUE\_SAR Reference Waveguide vG


This document shall not be reproduced, except in full or in part, without the written approval of MVG. The information contained herein is to be used only for the purpose for which it is submitted and is not to be released in whole or part without written approval of MVG.



## SAR REFERENCE WAVEGUIDE CALIBRATION REPORT

Ref: ACR.178.10.20.MVGB.A

### BODY SAR MEASUREMENT PLOTS @ 5800 MHz



Page: 12/13

*Template: ACR.DDD.N.YY.MVGB.ISSUE\_SAR Reference Waveguide vG*

*This document shall not be reproduced, except in full or in part, without the written approval of MVG. The information contained herein is to be used only for the purpose for which it is submitted and is not to be released in whole or part without written approval of MVG.*



## 8 LIST OF EQUIPMENT

| Equipment Summary Sheet            |                         |                    |                                               |                                               |
|------------------------------------|-------------------------|--------------------|-----------------------------------------------|-----------------------------------------------|
| Equipment Description              | Manufacturer / Model    | Identification No. | Current Calibration Date                      | Next Calibration Date                         |
| Flat Phantom                       | MVG                     | SN-13/09-SAM68     | Validated. No cal required.                   | Validated. No cal required.                   |
| COMOSAR Test Bench                 | Version 3               | NA                 | Validated. No cal required.                   | Validated. No cal required.                   |
| Network Analyzer                   | Rohde & Schwarz ZVM     | 100203             | 05/2019                                       | 05/2022                                       |
| Network Analyzer – Calibration kit | Rohde & Schwarz ZV-Z235 | 101223             | 05/2019                                       | 05/2022                                       |
| Calipers                           | Mitutoyo                | SN 0009732         | 10/2019                                       | 10/2022                                       |
| Reference Probe                    | MVG                     | EPGO333 SN 41/18   | 05/2020                                       | 05/2021                                       |
| Multimeter                         | Keithley 2000           | 1160271            | 02/2020                                       | 02/2023                                       |
| Signal Generator                   | Rohde & Schwarz SMB     | 106589             | 04/2019                                       | 04/2022                                       |
| Amplifier                          | Aethercomm              | SN 046             | Characterized prior to test. No cal required. | Characterized prior to test. No cal required. |
| Power Meter                        | NI-USB 5680             | 170100013          | 05/2019                                       | 05/2022                                       |
| Directional Coupler                | Narda 4216-20           | 01386              | Characterized prior to test. No cal required. | Characterized prior to test. No cal required. |
| Temperature and Humidity Sensor    | Control Company         | 150798832          | 11/2017                                       | 11/2020                                       |

Page: 13/13

Template: ACR.DDD.N.YY.MVGB.ISSUE\_SAR Reference Waveguide vG

This document shall not be reproduced, except in full or in part, without the written approval of MVG. The information contained herein is to be used only for the purpose for which it is submitted and is not to be released in whole or part without written approval of MVG.