

System Performance Check Report for CLA-13 - SN1015

Room Ambient Temperature: 23.0°C, Liquid Temperature: 22.0°C

Exposure Conditions

Frequency [MHz]	13.000	TSL Permittivity	56.6
Group / UID	CW / 0--	TSL Conductivity [S/m]	0.736
Conversion Factor	15.04	Phantom Section / TSL	Flat / Head Simulating Liquid

DASY Configuration

Probe Calibration Date	EX3DV4 - SN7313 2025-02-24	Phantom	ELI V8.0 (20deg probe tilt)
DAE Calibration Date	DAE4 Sn1591 2025-02-18	TSL Type	HBBL-600-10000
Software Version	16.4.0.5005		

Scans Setup

	Area Scan	Zoom Scan
Grid Extents [mm]	40.0 x 90.0	30.0 x 30.0 x 30.0
Grid Steps [mm]	10.0 x 15.0	6.0 x 6.0 x 1.5
Sensor Surface [mm]	3.0	1.4

Measurement Results

	Area Scan	Zoom Scan
psSAR1g [W/Kg]	0.056	0.054
psSAR10g [W/Kg]	0.045	0.033
Power Drift [dB]		0.01
Dist 3dB Peak [mm]		15.8
M2/M1 [%]		73.2

