



FCC ID Label Letter
attn: Reviewing Engineer
Federal Communications Commission
7435 Oakland Mills Road
Columbia, MD 21046

uCor V1.0.0

FCC ID: 2ABHFUCOR100

Jan. 1st, 2014

Subject: Kyma uCor 100 FCC ID Label

To Whom It May Concern:

The Kyma uCor V1.0.0 label is so small that it cannot include the statement specified in FCC Part 15 Sub-part A, Section 15.19 (a) (3). This device complies with Part 15 of the FCC Rules. Operation is subject to the following two conditions:

- (1) This device may not cause harmful interference and
- (2) This device must accept any interference received, including interference that may cause undesired operation."

Due to this reason, we are using the provisions of Part 15 Sub-part A, Section 15.19 (a) (5) and this statement appears in a prominent location in the user manual.

Please contact me if there is any information you may need.

Sincerely,

Moshe Mosesko

COO of Kyma Medical Technologies

A handwritten signature in blue ink, appearing to read "Moshe Mosesko".