

LABELING

Original Freigabeformular in TR archiviert.
Original form stored in the technical writing department.

Label inside of the device

FA. Nr. ##### - ##
SN YYMMCE####
Removing this label causes loss of warranty.
10000013740 04

MAX. IRRIGATION RATE
MAX. SPULLEISTUNG

2.0 l/min

MAX. PUMPING PRESSURE
MAX. PUMPENDRUCK

400 mmHg

MAINS VOLTAGE RANGE
NETZSPANNUNGSBEREICH

AC 100-240 V~

SUPPLY FREQUENCY
VERSORGUNGSFREQUENZ

50/60 Hz

POWER CONSUMPTION
LEISTUNGS-AUFNAHME

AC 100 V~ 0.65 A
AC 240 V~ 0.42 A

WARNING! Risk of fire. Replace fuse as marked.





Richard Wolf GmbH
Pforzheimer Straße 32,
75438 Knittlingen, Germany

CE 0124

REF REF Nr. SN YYMMCE#### YYYY-MM-DD

Name LAP SUCT./ IRRIG. PUMP

Supply mains 100-240 V~ 50/60 Hz 0.65 - 0.42 A

Classification I / IP41

T 3.15A H 250V

Contains FCC ID: 2AS5K-TSHW42
Contains IC: 25004-TSHW42A
HVIN: 1430

(01)#####
(11)YYMMDD(21)YYMMCE####

GS1 EAN
Data Matrix
DB 01/11/21

Issuing Agency:
GS1

Rear of device

Diagram of the rear of the device showing various components and their labels:

- 1109291: Quality Seal
- 1600041: Quality Seal
- 1601056: Component label
- 1600374: Component label
- 1601778: Component label
- 1109248: ONLY FOR SERVICE
- 1600385: Component label
- 10000016395 XX: Component label

Box, front

10000018931 XX

Diagram of the front of the box showing various symbols and labels:

- 1106671: Label
- PAP PE: Label
- Labels: Fragile, No open flame, Person with lightning bolt, 99%

Box, right side

10000013749 XX

10000009952 XX

Diagram of the right side of the box showing various symbols and labels:

- 10000022823 XX: Label

Revision history

Legend:		
Change request(s)		
Internal updates		
Rev.	WOM	ECR
Change		
01	2016	094
UDI; SN format; position of the symbols; format technical data; separate box label (10000018931 XX); classification BF -> I; Added: Name; date of manufacture; Removed: barcode (internal label); declaration of standards; 10000013745 XX + 10000013746 XX and included in box label;		
02	2017	393
Spelling REF (dot deleted) ; IEC 3.1		
03	2018	085
added: logistics label (10000022823); repositioned: test protocol label (10000009952), air out label (1109291)		
04	2020	066
added: symbol „Non-ionizing radiation“, check numbers; repositioned: label „Follow instructions for use“ (1601778)		

Product Management / Customer

Project Management (R&D)

Regulatory Affairs

WORLD OF MEDICINE / Approval

Overview device label P106 REF2216001 Wolf

W.O.M. Item number
10000013740 04

File FMT
10000013740 04.fmt
Name
TR

File CDR
10000013740 04.cdr
Date
23.11.2020