

PULSION Medical Systems SE | Hans-Riedl-Str. 21 |  
85622 Feldkirchen | Germany

SGS North America Inc.  
620 Old Peachtree Road  
SUITE 100  
Suwanee, Georgia  
United States

PULSION Medical Systems SE  
Thomas Leitner  
Hans-Riedl-Str. 21  
85622 Feldkirchen  
Germany

Phone: +49 89 45 99 14 - 453

Email: Thomas.leitner@getinge.com

Dear Sir/Madam,

**Re: Request for FCC module certification**

**FCC ID: FCC ID 2AVPQGC269016**

We hereby request a "Limited Single Modular" certification for the FCC ID referenced above.

The device meets the requirements indicated below as outlined in the FCC KDB 996369 D01  
Module Equip Auth Guide v01r04

- The radio element does not shield the radio frequency circuitry. The shielding is implemented in the terminal device; this is a limited single modular certification.
- The device has buffered modulation/data inputs to ensure that the device will comply with Part 15 requirements with any type of input signal;
- The device has power supply regulation on the module;
- The device has a permanently attached antenna, or contain a unique antenna connector, and be marketed and operated only with specific antenna(s), per Sections 15.203, 15.204(b), 15.204(c), 15.212(a), 2.929(b);
- The device has demonstrated compliance in a stand-alone configuration;
- The device will be labeled with its permanently affixed FCC ID label, or use an electronic display.
- The device does comply with all specific rules applicable to the transmitter including all the conditions provided in the integration instructions by the grantee;
- The device does comply with RF exposure requirements.

Sincerely,

Thomas Leitner

*Signature:*



**A. Single-modular transmitter** is a self-contained, physically delineated, component for which compliance can be demonstrated independent of the host operating conditions, and which complies with all eight requirements of Section 15.212(a)(1) as summarized below. See Section 15.212 for more detailed information, and Section 2.901 (and sub-sections that follow) for general certification requirements.

**B. Limited single-modular transmitter** is a transmitter that does not meet all eight requirements listed in Section 15.212(a) (1), and compliance can be demonstrated only for specific host and applicable operating conditions in which the transmitter will be used. For example, manufacturers have flexibility with respect to requirements such as module shielding, buffered modulation/data inputs and power supply regulation. If one or more of these functions (shielding, buffered modulation/data inputs and power supply regulation) are provided by a specific host or hosts, then the module can be granted as a limited module that is limited to that specific host or hosts. The responsible party must demonstrate how it will retain control over the final installation of the device, such that compliance of the product is ensured by limiting the installation to a specific host or hosts, for example.