

Sinocare Inc.
Request for Class II Permissive Change

FCC ID: 2AXH5001

Date: 2022-07-26

TO: Federal Communication Commission

Please be notified that we, the undersigned Sinocare Inc. declare that the reasons for this Class II permissive change are as below:

The difference between the Continuous Glucose Monitoring System (Model: i3) in original ID and current one is only on antenna gain. The antenna gain of the product has changed from 0.5dBi to 5.05dBi.

The original grant issued date is 11/15/2021

Thank you.

Sincerely.

Print Name: Fei Jiangfeng

Title: CGM Director

Signature:



On behalf of Company: Sinocare Inc.

Telephone: 8613521316621

E-mail: fei@sinocare.com