

Appendix A

Detailed System Check Results

1. System Performance Check

System Performance Check 2450 MHz Head
--



System Performance Check 2450MHz Head**D2450V2-SN 733**

Communication System: ; Frequency: 2450.000

Medium: Head Simulating Liquid. Medium parameters used: $f= 2450.000$ MHz; $\sigma= 1.79$ S/m; $\epsilon_r = 38.4$

DASY8 Configuration:

- Probe: EX3DV4 - SN7821; ConvF(7.5, 7.5, 7.5); Calibrated: 2024-08-29
- Sensor-Surface: 1.4 mm
- Electronics: DAE4ip Sn1830; Calibrated: 2024-10-18
- Phantom: Twin-SAM V8.0 (30deg probe tilt); Serial: 2146
- Measurement Software: cDASY8 V16.4.0.5005

Area Scan (40.0 mm x 96.0 mm): Measurement Grid: 10.0 mm x 12.0 mm

SAR (1g) = 12.2 W/kg; SAR (10g) = 5.82 W/kg;

Zoom Scan (30.0 mm x 30.0 mm x 30.0 mm): Measurement Grid: 5.0 mm x 5.0 mm x 1.5 mm

Power Drift = -0.00 dB

SAR (1g) = 12.4 W/kg; SAR (10g) = 5.78 W/kg;

M2/M1 [%]=78.0

Dist 3dB Peak [mm]=9.0

