

# RF System Check Test Report for CD1880:SN 1122

## Hardware Setup

Probe   Calibration Date	EF3DV3 - SN4028   September 04, 2023
DAE   Calibration Date	DAE4 Sn1546   March 11, 2024
Software Version	1.2.0.1713

## Communication Systems


Band Name	CD1880
Channel   Frequency [MHz]	50   1880.0
Communication Systems' Name	CW

## Grid Settings

Extent X [mm]	20.0	Step X [mm]	5.0
Extent Y [mm]	90.0	Step Y [mm]	5.0
Distance [mm]	15.0		

## Results

Emax [V/m]	<b>84.7</b>
Drift [dB]	<b>-0.02</b>



# RF System Check Test Report for CD2600:SN 1014

## Hardware Setup

Probe   Calibration Date	EF3DV3 - SN4028   September 04, 2023
DAE   Calibration Date	DAE4 Sn1546   March 11, 2024
Software Version	1.2.0.1713

## Communication Systems


Band Name	CD2600V3
Channel   Frequency [MHz]	50   2600.0
Communication Systems' Name	CW

## Grid Settings

Extent X [mm]	20.0	Step X [mm]	5.0
Extent Y [mm]	90.0	Step Y [mm]	5.0
Distance [mm]	15.0		

## Results

Emax [V/m]	81.7
Drift [dB]	0.02



# RF System Check Test Report for CD3500:SN 1006

## Hardware Setup

Probe   Calibration Date	EF3DV3 - SN4028   September 04, 2023
DAE   Calibration Date	DAE4 Sn1546   March 11, 2024
Software Version	1.2.0.1713

## Communication Systems

Band Name	CD3500V3
Channel   Frequency [MHz]	50   3500.0
Communication Systems' Name	CW

## Grid Settings

Extent X [mm]	20.0	Step X [mm]	5.0
Extent Y [mm]	60.0	Step Y [mm]	5.0
Distance [mm]	15.0		

## Results

Emax [V/m]	79.5
Drift [dB]	0.0

