

RF System Check Test Report for CD1880:SN 1212

Hardware Setup

Probe | Calibration Date EF3DV3 - SN4041 | March 04, 2025
DAE | Calibration Date DAE4ip Sn1621 | April 10, 2025
Software Version 1.2.6.2199

Communication Systems


Band Name CD1880
Channel | Frequency [MHz] 50 | 1880.0
Communication Systems' Name CW

Grid Settings

Extent X [mm]	20.0	Step X [mm]	5.0
Extent Y [mm]	90.0	Step Y [mm]	5.0
Distance [mm]	15.0		

Results

Emax [V/m] 86.8
Drift [dB] 0.06



RF System Check Test Report for CD2600:SN 1008

Hardware Setup

Probe | Calibration Date EF3DV3 - SN4041 | March 04, 2025
DAE | Calibration Date DAE4ip Sn1621 | April 10, 2025
Software Version 1.2.6.2199

Communication Systems


Band Name CD2600V3
Channel | Frequency [MHz] 50 | 2600.0
Communication Systems' Name CW

Grid Settings

Extent X [mm]	20.0	Step X [mm]	5.0
Extent Y [mm]	90.0	Step Y [mm]	5.0
Distance [mm]	15.0		

Results

Emax [V/m] **84.7**
Drift [dB] **-0.05**



RF System Check Test Report for CD3500:SN 1006

Hardware Setup

Probe | Calibration Date EF3DV3 - SN4041 | March 04, 2025
DAE | Calibration Date DAE4ip Sn1621 | April 10, 2025
Software Version 1.2.6.2199

Communication Systems


Band Name CD3500V3
Channel | Frequency [MHz] 50 | 3500.0
Communication Systems' Name CW

Grid Settings

Extent X [mm]	20.0	Step X [mm]	5.0
Extent Y [mm]	60.0	Step Y [mm]	5.0
Distance [mm]	15.0		

Results

Emax [V/m] 85.1
Drift [dB] 0.02



RF System Check Test Report for CD3500:SN 1006

Hardware Setup

Probe | Calibration Date EF3DV3 - SN4041 | March 04, 2025
DAE | Calibration Date DAE4ip Sn1621 | April 10, 2025
Software Version 1.2.6.2199

Communication Systems

Band Name CD3500V3
Channel | Frequency [MHz] 0 | 3700.0
Communication Systems' Name CW

Grid Settings

Extent X [mm]	20.0	Step X [mm]	5.0
Extent Y [mm]	60.0	Step Y [mm]	5.0
Distance [mm]	15.0		

Results

Emax [V/m] **77.6**
Drift [dB] **0.01**

