

System Check_Head_2450MHz

DUT: D2450V2-929

Communication System: CW; Frequency: 2450.0 MHz ; Duty Cycle: 1:1

Medium: HSL_2450_220812 Medium parameters used: $f = 2450.0$ MHz; $\sigma = 1.75$ S/m; $\epsilon_r = 38.3$

Ambient Temperature: 23.5°C; Liquid Temperature: 22.5°C

DASY6 Configuration:

- Probe: EX3DV4 - SN3976; ConvF(7.85, 7.85, 7.85); Calibrated: 2022-01-27
- Sensor-Surface: 1.4 mm
- Electronics: DAE4 Sn1707; Calibrated: 2022-01-12
- Phantom: ELI V8.0 (20deg probe tilt); Serial: 2153; Section: Flat
- Measurement Software: cDASY6 V6.6.0.13926
- UID: CW
- MAIA: Area Scan: N/A; Zoom Scan: N/A

Pin=50mW/Area Scan (40.0 mm x 96.0 mm): Measurement Grid: 10.0 mm x 12.0 mm


SAR (1g) = 2.53 W/kg; SAR (10g) = 1.20 W/kg;

Pin=50mW/Zoom Scan (30.0 mm x 30.0 mm x 30.0 mm): Measurement Grid: 5.0 mm x 5.0 mm x 1.5 mm

Power Drift = 0.11 dB

SAR (1g) = 2.52 W/kg; SAR (8g) = 1.28 W/kg; SAR (10g) = 1.16 W/kg;

psAPD (1.0cm², sq) = 25.2 [W/m²]; psAPD (4.0cm², sq) = 25.6 [W/m²]



System Check_Head_2450MHz

DUT: D2450V2-929

Communication System: CW; Frequency: 2450.0 MHz ; Duty Cycle: 1:1

Medium: HSL_2450_220819 Medium parameters used: $f = 2450.0$ MHz; $\sigma = 1.76$ S/m; $\epsilon_r = 38.4$

Ambient Temperature: 23.4°C; Liquid Temperature: 22.4°C

DASY6 Configuration:

- Probe: EX3DV4 - SN3728; ConvF(7.44, 7.44, 7.44); Calibrated: 2022-03-02
- Sensor-Surface: 1.4 mm
- Electronics: DAE4 Sn1647; Calibrated: 2022-01-19
- Phantom: ELI V8.0 (20deg probe tilt); Serial: 2157; Section: Flat
- Measurement Software: cDASY6 V6.6.0.13926
- UID: CW
- MAIA: Area Scan: Y; Zoom Scan: Y

Pin=50mW/Area Scan (40.0 mm x 96.0 mm): Measurement Grid: 10.0 mm x 12.0 mm


SAR (1g) = 2.53 W/kg; SAR (10g) = 1.20 W/kg;

Pin=50mW/Zoom Scan (30.0 mm x 30.0 mm x 30.0 mm): Measurement Grid: 5.0 mm x 5.0 mm x 1.5 mm

Power Drift = 0.09 dB

SAR (1g) = 2.42 W/kg; SAR (8g) = 1.28 W/kg; SAR (10g) = 1.12 W/kg;

psAPD (1.0cm², sq) = 24.2 [W/m²]; psAPD (4.0cm², sq) = 26.0 [W/m²]



System Check_Head_2450GHz

DUT: D2450V2-929

Communication System: CW; Frequency: 2450.0 MHz; Duty Cycle: 1:1

Medium: HSL_2450_221007 Medium parameters used: $f = 2450.0$ MHz; $\sigma = 1.82$ S/m; $\epsilon_r = 39.2$

Ambient Temperature: 23.5°C; Liquid Temperature: 22.5°C

DASY6 Configuration:

- Probe: EX3DV4 - SN7625; ConvF(7.91, 7.91, 7.91); Calibrated: 2022-01-27
- Sensor-Surface: 1.4 mm
- Electronics: DAE4 Sn1424; Calibrated: 2022-01-20
- Phantom: ELI V8.0 (20deg probe tilt); Serial: 2155; Section: Flat
- Measurement Software: 16.2.0.1425
- UID: CW, 0--

Pin=50mW/Area Scan (40.0 mm x 80.0 mm): Measurement Grid: 10.0 mm x 10.0 mm


SAR (1g) = 2.17 W/kg; SAR (10g) = 1.05 W/kg;

Pin=50mW/Zoom Scan (30.0 mm x 30.0 mm x 30.0 mm): Measurement Grid: 5.0 mm x 5.0 mm x 1.5 mm

Power Drift = 0.04 dB

SAR (1g) = 2.41 W/kg; SAR (8g) = 1.26 W/kg; SAR (10g) = 1.15 W/kg

psAPD (1.0cm², sq) = n/a [W/m²]; psAPD (4.0cm², sq) = n/a [W/m²]



System Check_Head_5250MHz

DUT: D5GHzV2-1128

Communication System: CW; Frequency: 5250.0 MHz ; Duty Cycle: 1:1

Medium: HSL_5G_220812 Medium parameters used: $f = 5250.0$ MHz; $\sigma = 4.61$ S/m; $\epsilon_r = 35.4$

Ambient Temperature: 23.5°C; Liquid Temperature: 22.5°C

DASY6 Configuration:

- Probe: EX3DV4 - SN3976; ConvF(5.47, 5.47, 5.47); Calibrated: 2022-01-27
- Sensor-Surface: 1.4 mm
- Electronics: DAE4 Sn1707; Calibrated: 2022-01-12
- Phantom: ELI V8.0 (20deg probe tilt); Serial: 2153; Section: Flat
- Measurement Software: cDASY6 V6.6.0.13926
- UID: CW
- MAIA: Area Scan: N/A; Zoom Scan: N/A

Pin=50mW/Area Scan (40.0 mm x 80.0 mm): Measurement Grid: 10.0 mm x 10.0 mm


SAR (1g) = 3.19 W/kg; SAR (10g) = 0.988 W/kg;

Pin=50mW/Zoom Scan (22.0 mm x 22.0 mm x 22.0 mm): Measurement Grid: 4.0 mm x 4.0 mm x 1.4 mm

Power Drift = 0.02 dB

SAR (1g) = 3.61 W/kg; SAR (8g) = 1.18 W/kg; SAR (10g) = 1.10 W/kg;

psAPD (1.0cm², sq) = 36.1 [W/m²]; psAPD (4.0cm², sq) = 23.6 [W/m²]



System Check_Head_5250GHz

DUT: D5GHzV2-1128

Communication System: CW; Frequency: 5250.0 MHz; Duty Cycle: 1:1

Medium: HSL_5G_221007 Medium parameters used: $f = 5250.0$ MHz; $\sigma = 4.71$ S/m; $\epsilon_r = 36.7$

Ambient Temperature: 23.5°C; Liquid Temperature: 22.5°C

DASY6 Configuration:

- Probe: EX3DV4 - SN7625; ConvF(5.45, 5.45, 5.45); Calibrated: 2022-01-27
- Sensor-Surface: 1.4 mm
- Electronics: DAE4 Sn1424; Calibrated: 2022-01-20
- Phantom: ELI V8.0 (20deg probe tilt); Serial: 2155; Section: Flat
- Measurement Software: 16.2.0.1425
- UID: CW, 0--

Pin=50mW/Area Scan (40.0 mm x 80.0 mm): Measurement Grid: 10.0 mm x 10.0 mm


SAR (1g) = 3.13 W/kg; SAR (10g) = 0.985 W/kg;

Pin=50mW/Zoom Scan (22.0 mm x 22.0 mm x 22.0 mm): Measurement Grid: 4.0 mm x 4.0 mm x 1.4 mm

Power Drift = -0.00 dB

SAR (1g) = 3.78 W/kg; SAR (8g) = 1.38 W/kg; SAR (10g) = 1.20 W/kg

psAPD (1.0cm², sq) = n/a [W/m²]; psAPD (4.0cm², sq) = n/a [W/m²]



System Check_Head_5600MHz

DUT: D5GHzV2-1128

Communication System: CW; Frequency: 5600.0 MHz; Duty Cycle: 1:1

Medium: HSL_5G_220812 Medium parameters used: $f = 5600.0$ MHz; $\sigma = 5.01$ S/m; $\epsilon_r = 34.8$

Ambient Temperature: 23.5°C; Liquid Temperature: 22.5°C

DASY6 Configuration:

- Probe: EX3DV4 - SN3976; ConvF(4.95, 4.95, 4.95); Calibrated: 2022-01-27
- Sensor-Surface: 1.4 mm
- Electronics: DAE4 Sn1707; Calibrated: 2022-01-12
- Phantom: ELI V8.0 (20deg probe tilt); Serial: 2153; Section: Flat
- Measurement Software: cDASY6 V6.6.0.13926
- UID: CW
- MAIA: Area Scan: N/A; Zoom Scan: N/A

Pin=50mW/Area Scan (40.0 mm x 80.0 mm): Measurement Grid: 10.0 mm x 10.0 mm


SAR (1g) = 3.62 W/kg; SAR (10g) = 1.11 W/kg;

Pin=50mW/Zoom Scan (22.0 mm x 22.0 mm x 22.0 mm): Measurement Grid: 4.0 mm x 4.0 mm x 1.4 mm

Power Drift = -0.01 dB

SAR (1g) = 4.08 W/kg; SAR (8g) = 1.33 W/kg; SAR (10g) = 1.13 W/kg;

psAPD (1.0cm², sq) = 40.8 [W/m²]; psAPD (4.0cm², sq) = 26.6 [W/m²]



System Check_Head_5600GHz

DUT: D5GHzV2-1128

Communication System: CW; Frequency: 5600.0 MHz; Duty Cycle: 1:1

Medium: HSL_5G_221008 Medium parameters used: $f = 5600.0$ MHz; $\sigma = 5.07$ S/m; $\epsilon_r = 36.2$

Ambient Temperature: 23.5°C; Liquid Temperature: 22.5°C

DASY6 Configuration:

- Probe: EX3DV4 - SN7625; ConvF(4.85, 4.85, 4.85); Calibrated: 2022-01-27
- Sensor-Surface: 1.4 mm
- Electronics: DAE4 Sn1424; Calibrated: 2022-01-20
- Phantom: ELI V8.0 (20deg probe tilt); Serial: 2155; Section: Flat
- Measurement Software: 16.2.0.1425
- UID: CW, 0--

Pin=50mW/Area Scan (40.0 mm x 80.0 mm): Measurement Grid: 10.0 mm x 10.0 mm


SAR (1g) = 3.39 W/kg; SAR (10g) = 1.05 W/kg;

Pin=50mW/Zoom Scan (22.0 mm x 22.0 mm x 22.0 mm): Measurement Grid: 4.0 mm x 4.0 mm x 1.4 mm

Power Drift = 0.00 dB

SAR (1g) = 4.08 W/kg; SAR (8g) = 1.49 W/kg; SAR (10g) = 1.29 W/kg

psAPD (1.0cm², sq) = n/a [W/m²]; psAPD (4.0cm², sq) = n/a [W/m²]



System Check_Head_5750MHz

DUT: D5GHzV2-1128

Communication System: CW; Frequency: 5750.0 MHz ; Duty Cycle: 1:1

Medium: HSL_5G_220812 Medium parameters used: $f = 5750.0$ MHz; $\sigma = 5.19$ S/m; $\epsilon_r = 34.5$

Ambient Temperature: 23.5°C; Liquid Temperature: 22.5°C

DASY6 Configuration:

- Probe: EX3DV4 - SN3976; ConvF(4.87, 4.87, 4.87); Calibrated: 2022-01-27
- Sensor-Surface: 1.4 mm
- Electronics: DAE4 Sn1707; Calibrated: 2022-01-12
- Phantom: ELI V8.0 (20deg probe tilt); Serial: 2153; Section: Flat
- Measurement Software: cDASY6 V6.6.0.13926
- UID: CW
- MAIA: Area Scan: N/A; Zoom Scan: N/A

Pin=50mW/Area Scan (40.0 mm x 80.0 mm): Measurement Grid: 10.0 mm x 10.0 mm


SAR (1g) = 3.42 W/kg; SAR (10g) = 1.05 W/kg;

Pin=50mW/Zoom Scan (22.0 mm x 22.0 mm x 22.0 mm): Measurement Grid: 4.0 mm x 4.0 mm x 1.4 mm

Power Drift = -0.02 dB

SAR (1g) = 3.86 W/kg; SAR (8g) = 1.25 W/kg; SAR (10g) = 1.06 W/kg;

psAPD (1.0cm², sq) = 38.6 [W/m²]; psAPD (4.0cm², sq) = 25.0 [W/m²]



System Check_Head_5750GHz

DUT: D5GHzV2-1128

Communication System: CW; Frequency: 5750.0 MHz; Duty Cycle: 1:1

Medium: HSL_5G_221008 Medium parameters used: $f = 5750.0$ MHz; $\sigma = 5.23$ S/m; $\epsilon_r = 36.0$

Ambient Temperature: 23.5°C; Liquid Temperature: 22.5°C

DASY6 Configuration:

- Probe: EX3DV4 - SN7625; ConvF(4.93, 4.93, 4.93); Calibrated: 2022-01-27
- Sensor-Surface: 1.4 mm
- Electronics: DAE4 Sn1424; Calibrated: 2022-01-20
- Phantom: ELI V8.0 (20deg probe tilt); Serial: 2155; Section: Flat
- Measurement Software: 16.2.0.1425
- UID: CW, 0--

Pin=50mW/Area Scan (40.0 mm x 80.0 mm): Measurement Grid: 10.0 mm x 10.0 mm


SAR (1g) = 3.25 W/kg; SAR (10g) = 1.01 W/kg;

Pin=50mW/Zoom Scan (22.0 mm x 22.0 mm x 22.0 mm): Measurement Grid: 4.0 mm x 4.0 mm x 1.4 mm

Power Drift = 0.03 dB

SAR (1g) = 3.93 W/kg; SAR (8g) = 1.43 W/kg; SAR (10g) = 1.24 W/kg

psAPD (1.0cm², sq) = n/a [W/m²]; psAPD (4.0cm², sq) = n/a [W/m²]



System Check_Head_6500MHz

DUT: D6.5GHzV2-1003

Communication System: CW; Frequency: 6500.0 MHz; Duty Cycle: 1:1

Medium: HSL_6G_220812 Medium parameters used: $f = 6500.0$ MHz; $\sigma = 6.06$ S/m; $\epsilon_r = 34.42$

Ambient Temperature: 23.5°C; Liquid Temperature: 22.5°C

DASY6 Configuration:

- Probe: EX3DV4 - SN3976; ConvF(5.8, 5.8, 5.8); Calibrated: 2022-01-27
- Sensor-Surface: 1.4 mm
- Electronics: DAE4 Sn1707; Calibrated: 2022-01-12
- Phantom: ELI V8.0 (20deg probe tilt); Serial: 2153; Section: Flat
- Measurement Software: cDASY6 V6.6.0.13926
- UID: CW
- MAIA: Area Scan: N/A; Zoom Scan: N/A

Pin=100mW/Area Scan (51.0 mm x 51.0 mm): Measurement Grid: 8.5 mm x 8.5 mm


SAR (1g) = 21.4 W/kg; SAR (10g) = 4.32 W/kg;

Pin=100mW/Zoom Scan (23.8 mm x 23.8 mm x 22.0 mm): Measurement Grid: 3.4 mm x 3.4 mm x 1.4 mm

Power Drift = 0.01 dB

SAR (1g) = 26.4 W/kg; SAR (8g) = 5.88 W/kg; SAR (10g) = 5.05 W/kg;

psAPD (1.0cm², sq) = 264 [W/m²]; psAPD (4.0cm², sq) = 119



System Check_Head_6500MHz

DUT: D6.5GHzV2-1003

Communication System: CW; Frequency: 6500.0 MHz; Duty Cycle: 1:1

Medium: HSL_6G_221006 Medium parameters used: $f = 6500.0$ MHz; $\sigma = 6.18$ S/m; $\epsilon_r = 34.67$

Ambient Temperature: 23.5°C; Liquid Temperature: 22.5°C

DASY6 Configuration:

- Probe: EX3DV4 - SN7625; ConvF(5.35, 5.35, 5.35); Calibrated: 2022-01-27
- Sensor-Surface: 1.4 mm
- Electronics: DAE4 Sn1424; Calibrated: 2022-01-20
- Phantom: ELI V8.0 (20deg probe tilt); Serial: 2155; Section: Flat
- Measurement Software: 16.2.0.1425
- UID: CW
- MAIA: Area Scan: N/A; Zoom Scan: N/A

Pin=100mW/Area Scan (51.0 mm x 51.0 mm): Measurement Grid: 8.5 mm x 8.5 mm


SAR (1g) = 21.5 W/kg; SAR (10g) = 4.33 W/kg;

Pin=100mW/Zoom Scan (23.8 mm x 23.8 mm x 22.0 mm): Measurement Grid: 3.4 mm x 3.4 mm x 1.4 mm

Power Drift = 0.03 dB

SAR (1g) = 26.8 W/kg; SAR (8g) = 5.89 W/kg; SAR (10g) = 5.10 W/kg;

psAPD (1.0cm², sq) = 268 [W/m²]; psAPD (4.0cm², sq) = 120



System Check_Head_6500MHz

DUT: D6.5GHzV2-1003

Communication System: CW; Frequency: 6500.0 MHz; Duty Cycle: 1:1

Medium: HSL_6G_221008 Medium parameters used: $f = 6500.0$ MHz; $\sigma = 6.13$ S/m; $\epsilon_r = 34.69$

Ambient Temperature: 23.5°C; Liquid Temperature: 22.5°C

DASY6 Configuration:

- Probe: EX3DV4 - SN7625; ConvF(5.35, 5.35, 5.35); Calibrated: 2022-01-27
- Sensor-Surface: 1.4 mm
- Electronics: DAE4 Sn1424; Calibrated: 2022-01-20
- Phantom: ELI V8.0 (20deg probe tilt); Serial: 2155; Section: Flat
- Measurement Software: cDASY6 V6.6.0.13926
- UID: CW
- MAIA: Area Scan: N/A; Zoom Scan: N/A

Pin=100mW/Area Scan (51.0 mm x 51.0 mm): Measurement Grid: 8.5 mm x 8.5 mm

SAR (1g) = 21.4 W/kg; SAR (10g) = 4.34 W/kg;

Pin=100mW/Zoom Scan (23.8 mm x 23.8 mm x 22.0 mm): Measurement Grid: 3.4 mm x 3.4 mm x 1.4 mm

Power Drift = 0.03 dB

SAR (1g) = 26.3 W/kg; SAR (8g) = 5.90 W/kg; SAR (10g) = 5.06 W/kg;

psAPD (1.0cm², sq) = 261 [W/m²]; psAPD (4.0cm², sq) = 119

