

## Ken Meyer

---

**From:** oetech@fccsun27w.fcc.gov  
**Sent:** Monday, July 12, 2010 11:58 AM  
**To:** Ken Meyer  
**Subject:** Response to Inquiry to FCC (Tracking Number 775017)



### Office of Engineering and Technology

#### Inquiry:

---Reply from Customer on 07/08/2010---

As stated in the previous response, the lower limit of our printing process is a font size of 4, so the font size can not be reduced. To fit the FCC ID information on the device would require a font size of 2. At a font size of 2, the FCC ID would become an illegible smear, even under magnification.

I understand the importance of the FCC ID being on the device. And I understand the importance of being conservative regarding granting waivers. However, given the extremely small size of our device, this situation seems to be exactly fitting to the intent of 2.925(f).

#### **Response:**

We agree your proposal to put the FCC ID in the user manual is consistent with the two precedents you included in your inquiry. You may place the FCC ID in the user's manual in this case since the ID won't fit on the device. Please include a copy of this response with your application package for the TCB grant.

[hearing aid label control drawing](#)

[Grant of Eq Authorization for competitive aid](#)

[Grant of Eq Authorization for a second competitive aid](#)

[Figure 1 showing curved nature and relative size of hearing aid](#)

[Figure 2 showing available printable area and FCC ID](#)

Do not reply to this message. Please select the [Reply to an Inquiry Response](#) link from the OET Inquiry System to add any additional information pertaining to this inquiry.