



Notes:

1. Material: 0.010 Velvet Polycarbonate
Adhesive: MIS-01-022-02
Finish: Laminate white barrier layer for opacity
(General Formulation PPS2W-STC-OL/53#
or equivalent.)
2. Text Copy: Mortara Black per MIS-01-021-05.
Background: Mortara White per MIS-01-021-03.
3. The label shall be a "UL" recognized marking and labeling system, for use on various metals, painted surfaces, and plastic, to indicate conformance to this requirement, the supplier shall identify conformance with appropriate markings on direct packaging.

4. Lines depicting label borders are for reference only and should not appear on finished copy.

Measurements in Inches
Tolerance: +/- .006 unless otherwise noted.

"ELECTRONIC ORIGINAL" Any Changes To This Document Done Manually Void The Data Base Original.		CONFIDENTIAL This print and the design therein represented are the property of Mortara Instrument, Incorporated, and may not be used in any manner detrimental to its interest. This print has been issued with the understanding that it will not be reproduced, that no disclosure of its contents will be made to others, that it will not be used for manufacture for others, and will be returned upon request.	Mortara INSTRUMENT 7865 N. 86th STREET, MILWAUKEE, WI 53224 TEL. 414-354-1600	
DRAWN BY: Diane A. Cole	DATE: 10-07-07		TITLE: LABEL X12+ NAMEPLATE 915 MHz	
DOCUMENTATION CHECKED BY:	DATE:		LATEST DOCUMENT CHANGE:	
ENGINEERING APPROVED BY:	DATE:		DR: 1	
MANUFACTURING APPROVED BY:	DATE:		ECO: 6282	
Q.A. APPROVED BY:	DATE:		DATE: 10-27-07	
OTHER:	DATE:		BY: Diane A. Cole	
		SCALE: FULL	PART NUMBER: 9050 - 057 - 04	REV: A1
		SHEET 1 OF 2		

Mortara INSTRUMENT Milwaukee, WI U.S.A. **Model: X12+915**

This device complies with part 15 of the FCC rules. Operation is subject to the following two conditions: (1) this device may not cause harmful interference, and (2) this device must accept any interference received, including interference that may cause undesired operation. FCC ID: HJR-X12P-915 IC: 3758B-X12P915



MEDICAL EQUIPMENT
UL2601-1, IEC60601-1
CAN/CSA C22.2 No. 601.1
IEC60601-2-25

PAGE:	SCALE :	PART NUMBER:	REV:
2 of 2	NTS	9050 - 057 - 04	A1