



APPENDIX E - MEASURED FLUID DIELECTRIC PARAMETERS

Title

SubTitle

November 21, 2013 06:56 PM

Frequency	e'	e''
800.000000 MI	53.4739	20.1925
803.000000 MI	53.4289	20.1918
806.000000 MI	53.3889	20.2009
809.000000 MI	53.3349	20.1701
812.000000 MI	53.3119	20.2121
815.000000 MI	53.2849	20.1774
818.000000 MI	53.2579	20.1875
821.000000 MI	53.1999	20.2031
824.000000 MI	53.1669	20.1932
827.000000 MI	53.1489	20.2109
830.000000 MI	53.1219	20.2154
833.000000 MI	53.0949	20.2198
836.000000 MI	53.0749	20.2272
839.000000 MI	53.0379	20.2462
842.000000 MI	53.0009	20.2213
845.000000 MI	52.9809	20.2395
848.000000 MI	52.9289	20.2150
851.000000 MI	52.9019	20.2252
854.000000 MI	52.9089	20.2117
857.000000 MI	52.8879	20.2086
860.000000 MI	52.8569	20.2089
863.000000 MI	52.8299	20.1868
866.000000 MI	52.8159	20.1706
869.000000 MI	52.7959	20.1672
872.000000 MI	52.7789	20.1416
875.000000 MI	52.7419	20.0917
878.000000 MI	52.7479	20.0996
881.000000 MI	52.7109	20.1029
884.000000 MI	52.6749	20.0532
887.000000 MI	52.6929	20.0355
890.000000 MI	52.6349	20.0364
893.000000 MI	52.6279	19.9920
896.000000 MI	52.5929	19.9689
899.000000 MI	52.5869	19.9399
902.000000 MI	52.5689	19.9314

905.000000 MI	52.540%	19.8852
908.000000 MI	52.515%	19.8997
911.000000 MI	52.489%	19.8839
914.000000 MI	52.474%	19.8796
917.000000 MI	52.451%	19.8575
920.000000 MI	52.390%	19.8404
923.000000 MI	52.384%	19.8325
926.000000 MI	52.367%	19.8489
929.000000 MI	52.349%	19.8355
932.000000 MI	52.307%	19.8619
935.000000 MI	52.271%	19.8509
938.000000 MI	52.261%	19.8294
941.000000 MI	52.238%	19.8633
944.000000 MI	52.188%	19.8529
947.000000 MI	52.151%	19.8623
950.000000 MI	52.116%	19.8758

Title

SubTitle

June 25, 2014 04:28 PM

Frequency	e'	e''
1.750000000 GHz	54.7671	14.8314
1.752000000 GHz	54.8901	14.8636
1.754000000 GHz	54.9621	14.8838
1.756000000 GHz	54.9761	14.8818
1.758000000 GHz	54.8781	14.8680
1.760000000 GHz	54.7681	14.8310
1.762000000 GHz	54.6841	14.8099
1.764000000 GHz	54.6801	14.8045
1.766000000 GHz	54.7461	14.8440
1.768000000 GHz	54.8541	14.8581
1.770000000 GHz	54.9301	14.9000
1.772000000 GHz	54.8681	14.9048
1.774000000 GHz	54.7791	14.8695
1.776000000 GHz	54.6881	14.8282
1.778000000 GHz	54.6031	14.8153
1.780000000 GHz	54.6411	14.8383
1.782000000 GHz	54.7041	14.8609
1.784000000 GHz	54.8211	14.9106
1.786000000 GHz	54.8651	14.9197
1.788000000 GHz	54.8121	14.8950
1.790000000 GHz	54.6981	14.8865
1.792000000 GHz	54.6081	14.8714
1.794000000 GHz	54.5821	14.8602
1.796000000 GHz	54.6051	14.8877
1.798000000 GHz	54.7031	14.9235
1.800000000 GHz	54.7971	14.9628
1.802000000 GHz	54.7981	14.9729
1.804000000 GHz	54.7321	14.9594
1.806000000 GHz	54.6121	14.9271
1.808000000 GHz	54.5391	14.9130
1.810000000 GHz	54.5401	14.9229
1.812000000 GHz	54.5951	14.9517
1.814000000 GHz	54.6661	14.9890
1.816000000 GHz	54.7291	15.0139
1.818000000 GHz	54.7401	15.0261

1.820000000 Gf	54.622	15.0081
1.822000000 Gf	54.540	14.9838
1.824000000 Gf	54.484	14.9590
1.826000000 Gf	54.476	14.9799
1.828000000 Gf	54.563	15.0134
1.830000000 Gf	54.630	15.0458
1.832000000 Gf	54.683	15.0872
1.834000000 Gf	54.634	15.0413
1.836000000 Gf	54.558	15.0195
1.838000000 Gf	54.462	15.0204
1.840000000 Gf	54.435	15.0067
1.842000000 Gf	54.468	15.0292
1.844000000 Gf	54.535	15.0570
1.846000000 Gf	54.626	15.0818
1.848000000 Gf	54.628	15.0949
1.850000000 Gf	54.564	15.0739



APPENDIX F – PHANTOM CERTIFICATE OF CONFORMITY

Certificate of conformity / First Article Inspection

Item	SAM Twin Phantom V4.0
Type No	QD 000 P40 C
Series No	TP-1150 and higher
Manufacturer / Origin	Untersee Composites Hauptstr. 69 CH-8559 Fruthwilen Switzerland

Tests

The series production process used allows the limitation to test of first articles.

Complete tests were made on the pre-series Type No. QD 000 P40 AA, Serial No. TP-1001 and on the series first article Type No. QD 000 P40 BA, Serial No. TP-1006. Certain parameters have been retested using further series units (called samples).

Test	Requirement	Details	Units tested
Shape	Compliance with the geometry according to the CAD model.	IT'IS CAD File (*)	First article, Samples
Material thickness	Compliant with the requirements according to the standards	2mm +/- 0.2mm in specific areas; 6mm +/- 0.2mm at ERP	First article, Samples
Material parameters	Dielectric parameters for required frequencies	200 MHz – 3 GHz Relative permittivity < 5 Loss tangent < 0.05.	Material sample TP 104-5
Material resistivity	The material has been tested to be compatible with the liquids defined in the standards if handled and cleaned according to the instructions	DEGMBE based simulating liquids	Pre-series, First article, Samples

Standards

[1] CENELEC EN 50361

[2] IEEE Std 1528-200x Draft CD 1.1 (Dec 02)

[3] IEC 62209/CD (Nov 02)

(*) The IT'IS CAD file is derived from [2] and is also within the tolerance requirements of the shapes of [1] and [3].

Conformity

Based on the sample tests above, we certify that this item is in compliance with the uncertainty requirements of SAR measurements specified in standard [1] and draft standards [2] and [3].

Date

7.8.2003

Signature / Stamp

s p e a g

Schmid & Partner Engineering AG
 Zeughausstrasse 43, 8004 Zurich, Switzerland
 Phone +41 1 245 9700, Fax +41 1 245 9779
 info@speag.com, http://www.speag.com



APPENDIX G - DAE CALIBRATION CERTIFICATE



Accredited by the Swiss Accreditation Service (SAS)

Accreditation No.: **SCS 108**

The Swiss Accreditation Service is one of the signatories to the EA
Multilateral Agreement for the recognition of calibration certificates

Client **MET Laboratories**

Certificate No: **DAE3-584_Jul13**

CALIBRATION CERTIFICATE

Object **DAE3 - SD 000 D03 AA - SN: 584**

Calibration procedure(s) **QA CAL-06.v26**
Calibration procedure for the data acquisition electronics (DAE)

Calibration date: **July 18, 2013**

This calibration certificate documents the traceability to national standards, which realize the physical units of measurements (SI).
The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.

All calibrations have been conducted in the closed laboratory facility: environment temperature (22 ± 3)°C and humidity < 70%.

Calibration Equipment used (M&TE critical for calibration)

Primary Standards	ID #	Cal Date (Certificate No.)	Scheduled Calibration
Keithley Multimeter Type 2001	SN: 0810278	02-Oct-12 (No:12728)	Oct-13
Secondary Standards	ID #	Check Date (in house)	Scheduled Check
Auto DAE Calibration Unit	SE UWS 053 AA 1001	07-Jan-13 (in house check)	In house check: Jan-14
Calibrator Box V2.1	SE UMS 006 AA 1002	07-Jan-13 (in house check)	In house check: Jan-14

Calibrated by:	Name	Function	Signature
	Dominique Steffen	Technician	
Approved by:	Fin Bomholt	Deputy Technical Manager	

Issued: July 18, 2013

This calibration certificate shall not be reproduced except in full without written approval of the laboratory.



Accredited by the Swiss Accreditation Service (SAS)

The Swiss Accreditation Service is one of the signatories to the EA
Multilateral Agreement for the recognition of calibration certificates

Accreditation No.: **SCS 108**

Glossary

DAE	data acquisition electronics
Connector angle	information used in DASY system to align probe sensor X to the robot coordinate system.

Methods Applied and Interpretation of Parameters

- *DC Voltage Measurement:* Calibration Factor assessed for use in DASY system by comparison with a calibrated instrument traceable to national standards. The figure given corresponds to the full scale range of the voltmeter in the respective range.
- *Connector angle:* The angle of the connector is assessed measuring the angle mechanically by a tool inserted. Uncertainty is not required.
- The following parameters as documented in the Appendix contain technical information as a result from the performance test and require no uncertainty.
 - *DC Voltage Measurement Linearity:* Verification of the Linearity at +10% and -10% of the nominal calibration voltage. Influence of offset voltage is included in this measurement.
 - *Common mode sensitivity:* Influence of a positive or negative common mode voltage on the differential measurement.
 - *Channel separation:* Influence of a voltage on the neighbor channels not subject to an input voltage.
 - *AD Converter Values with inputs shorted:* Values on the internal AD converter corresponding to zero input voltage
 - *Input Offset Measurement:* Output voltage and statistical results over a large number of zero voltage measurements.
 - *Input Offset Current:* Typical value for information; Maximum channel input offset current, not considering the input resistance.
 - *Input resistance:* Typical value for information: DAE input resistance at the connector, during internal auto-zeroing and during measurement.
 - *Low Battery Alarm Voltage:* Typical value for information. Below this voltage, a battery alarm signal is generated.
 - *Power consumption:* Typical value for information. Supply currents in various operating modes.

DC Voltage Measurement

A/D - Converter Resolution nominal

High Range: 1LSB = 6.1 μ V , full range = -100...+300 mV

Low Range: 1LSB = 61nV , full range = -1.....+3mV

DASY measurement parameters: Auto Zero Time: 3 sec; Measuring time: 3 sec

Calibration Factors	X	Y	Z
High Range	404.662 $\pm$ 0.02% (k=2)	404.844 $\pm$ 0.02% (k=2)	404.267 $\pm$ 0.02% (k=2)
Low Range	3.93273 $\pm$ 1.50% (k=2)	3.98110 $\pm$ 1.50% (k=2)	3.97938 $\pm$ 1.50% (k=2)

Connector Angle

Connector Angle to be used in DASY system	141.0 ° $\pm$ 1 °
---	-------------------

Appendix

1. DC Voltage Linearity

High Range	Reading (μV)	Difference (μV)	Error (%)
Channel X + Input	199996.51	-1.40	-0.00
Channel X + Input	20001.72	1.16	0.01
Channel X - Input	-19998.74	2.23	-0.01
Channel Y + Input	199995.49	-2.08	-0.00
Channel Y + Input	20001.55	1.15	0.01
Channel Y - Input	-19997.38	3.75	-0.02
Channel Z + Input	199996.10	-1.20	-0.00
Channel Z + Input	19999.08	-1.38	-0.01
Channel Z - Input	-19999.72	1.40	-0.01

Low Range	Reading (μV)	Difference (μV)	Error (%)
Channel X + Input	2002.82	1.92	0.10
Channel X + Input	202.17	0.93	0.46
Channel X - Input	-198.65	-0.18	0.09
Channel Y + Input	2002.22	1.54	0.08
Channel Y + Input	200.42	-0.71	-0.35
Channel Y - Input	-199.73	-1.09	0.55
Channel Z + Input	2000.94	0.28	0.01
Channel Z + Input	200.01	-1.13	-0.56
Channel Z - Input	-200.29	-1.60	0.80

2. Common mode sensitivity

DASY measurement parameters: Auto Zero Time: 3 sec; Measuring time: 3 sec

	Common mode Input Voltage (mV)	High Range Average Reading (μV)	Low Range Average Reading (μV)
Channel X	200	3.04	1.31
	- 200	-0.45	-1.35
Channel Y	200	3.69	2.42
	- 200	-3.59	-3.54
Channel Z	200	-7.37	-7.51
	- 200	5.30	5.09

3. Channel separation

DASY measurement parameters: Auto Zero Time: 3 sec; Measuring time: 3 sec

	Input Voltage (mV)	Channel X (μV)	Channel Y (μV)	Channel Z (μV)
Channel X	200	-	-0.18	-3.67
Channel Y	200	6.93	-	-0.17
Channel Z	200	7.39	5.46	-

IMPORTANT NOTICE

USAGE OF THE DAE 3

The DAE unit is a delicate, high precision instrument and requires careful treatment by the user. There are no serviceable parts inside the DAE. Special attention shall be given to the following points:

Battery Exchange: The battery cover of the DAE3 unit is connected to a fragile 3-pin battery connector. Customer is responsible to apply outmost caution not to bend or damage the connector when changing batteries.

Shipping of the DAE: Before shipping the DAE to SPEAG for calibration the customer shall remove the batteries and pack the DAE in an antistatic bag. This antistatic bag shall then be packed into a larger box or container which protects the DAE from impacts transportation. The package shall be marked to indicate that a fragile instrument is inside.

E-Stop Failures: Touch detection may be malfunctioning due to broken magnets in the E-stop. Rough handling of the E-stop may lead to damage of these magnets. Touch and collision errors are often caused by dust and dirt accumulated in the E-stop. To prevent E-stop failure, Customer shall always mount the probe to the DAE carefully and keep the DAE unit in a non-dusty environment if not used for measurements.

Repair: Minor repairs are performed at no extra cost during the annual calibration. However, SPEAG reserves the right to charge for any repair especially if rough unprofessional handling caused the defect.

DASY Configuration Files: Since the exact values of the DAE input resistances, as measured during the calibration procedure of a DAE unit, are not used by the DASY software, a nominal value of 200 MOhm is given in the corresponding configuration file.

Important Note:

Warranty and calibration is void if the DAE unit is disassembled partly or fully by the Customer.

Important Note:

Never attempt to grease or oil the E-stop assembly. Cleaning and readjusting of the E-stop assembly is allowed by certified SPEAG personnel only and is part of the annual calibration procedure.

Important Note:

To prevent damage of the DAE probe connector pins, use great care when installing the probe to the DAE. Carefully connect the probe with the connector notch oriented in the mating position. Avoid any rotational movement of the probe body versus the DAE while turning the locking nut of the connector. The same care shall be used when disconnecting the probe from the DAE.

4. AD-Converter Values with inputs shorted

DASY measurement parameters: Auto Zero Time: 3 sec; Measuring time: 3 sec

	High Range (LSB)	Low Range (LSB)
Channel X	16109	15852
Channel Y	16158	16754
Channel Z	16279	16611

5. Input Offset Measurement

DASY measurement parameters: Auto Zero Time: 3 sec; Measuring time: 3 sec

Input 10M Ω

	Average (μ V)	min. Offset (μ V)	max. Offset (μ V)	Std. Deviation (μ V)
Channel X	1.30	-0.50	3.30	0.53
Channel Y	-0.18	-1.65	1.35	0.50
Channel Z	-0.26	-1.48	1.14	0.47

6. Input Offset Current

Nominal Input circuitry offset current on all channels: <25fA

7. Input Resistance (Typical values for information)

	Zeroing (kOhm)	Measuring (MOhm)
Channel X	200	200
Channel Y	200	200
Channel Z	200	200

8. Low Battery Alarm Voltage (Typical values for information)

Typical values	Alarm Level (VDC)
Supply (+ Vcc)	+7.9
Supply (- Vcc)	-7.6

9. Power Consumption (Typical values for information)

Typical values	Switched off (mA)	Stand by (mA)	Transmitting (mA)
Supply (+ Vcc)	+0.01	+6	+14
Supply (- Vcc)	-0.01	-8	-9