

DFS Portions of FCC CFR47 PART 15 SUBPART E and INDUSTRY CANADA RSS-210 ISSUE 7

CERTIFICATION TEST REPORT

FOR

802.11 A/B/G WIRELESS RADIO

MODEL NUMBER: WELCH ALLYN 802.11 A/B/G WIRELESS RADIO

FCC ID: PGUWA11ABG09 IC: 4168A-WA11ABG09

REPORT NUMBER: 09U12573-1

ISSUE DATE: MAY 12, 2009

Prepared for

WELCH ALLYN PROTOCOL, INC. 8500 SW CREEKSIDE PLACE BEAVERTON OR., 97008-7107, U.S.A.

Prepared by

COMPLIANCE CERTIFICATION SERVICES 47173 BENICIA STREET FREMONT, CA 94538, U.S.A.

TEL: (510) 771-1000 FAX: (510) 661-0888



REPORT NO: 09U12573-1 FCC ID: PGUWA11ABG09 DATE: MAY 12, 2009 IC: 4168A-WA11ABG09

Revision History

Rev.	Issue Date	Revisions	Revised By
	05/12/09	Initial Issue	M.Heckrotte

TABLE OF CONTENTS

1.	. ATTESTATION OF TEST RESULTS	4
2.	. TEST METHODOLOGY	5
	FACILITIES AND ACCREDITATION	
	. CALIBRATION AND UNCERTAINTY	
	4.1. MEASURING INSTRUMENT CALIBRATION	_
	4.2. MEASUREMENT UNCERTAINTY	5
5.	DYNAMIC FREQUENCY SELECTION	6
	5.1. OVERVIEW	
	5.1.2. TEST AND MEASUREMENT SYSTEM 5.1.3. SETUP OF EUT	9
	5.1.4. DESCRIPTION OF EUT	13
	5.2. RESULTS FOR 20 MHz BANDWIDTH	14
	5.2.2. PLOTS OF RADAR WAVEFORM AND WLAN TRAFFICE 5.2.3. MOVE AND CLOSING TIME	16
	5.2.4. NON-OCCUPANCY	21
•	CETUD DUOTOS	22

DATE: MAY 12, 2009 IC: 4168A-WA11ABG09

1. ATTESTATION OF TEST RESULTS

COMPANY NAME: WELCH ALLYN PROTOCOL, INC.

8500 SW CREEKSIDE PLACE

BEAVERTON

OR., 97008-7107, U.S.A.

EUT DESCRIPTION: 802.11 A/B/G WIRELESS RADIO

MODEL: WELCH ALLYN 802.11 A/B/G WIRELESS RADIO

SERIAL NUMBER: 02289

DATE TESTED: MAY 11 TO 12, 2009

APPLICABLE STANDARDS

STANDARD TEST RESULTS

DFS Portion of CFR 47 Part 15 Subpart E

Pass

DFS Portion of INDUSTRY CANADA RSS-210 Issue 7 Annex 9

Pass

Compliance Certification Services, Inc. (CCS) tested the above equipment in accordance with the requirements set forth in the above standards. All indications of Pass/Fail in this report are opinions expressed by CCS based on interpretations and/or observations of test results. The test results show that the equipment tested is capable of demonstrating compliance with the requirements as documented in this report.

Note: The results documented in this report apply only to the tested sample, under the conditions and modes of operation as described herein. This document may not be altered or revised in any way unless done so by CCS and all revisions are duly noted in the revisions section. Any alteration of this document not carried out by CCS will constitute fraud and shall nullify the document. No part of this report may be used to claim product certification, approval, or endorsement by NVLAP, NIST, or any government agency.

Approved & Released For CCS By:

Tested By:

MICHAEL HECKROTTE DIRECTOR OF ENGINEERING

MH

COMPLIANCE CERTIFICATION SERVICES

DOUGLAS ANDERSON EMC ENGINEER

Douglas Combuser.

COMPLIANCE CERTIFICATION SERVICES

2. TEST METHODOLOGY

The tests documented in this report were performed in accordance with FCC CFR 47 Part 15, FCC 06-96 and RSS-210 Issue 7.

3. FACILITIES AND ACCREDITATION

The test sites and measurement facilities used to collect data are located at 47173 Benicia Street, Fremont, California, USA.

CCS is accredited by NVLAP, Laboratory Code 200065-0. The full scope of accreditation can be viewed at http://www.ccsemc.com.

4. CALIBRATION AND UNCERTAINTY

4.1. MEASURING INSTRUMENT CALIBRATION

The measuring equipment utilized to perform the tests documented in this report has been calibrated in accordance with the manufacturer's recommendations, and is traceable to recognized national standards.

4.2. MEASUREMENT UNCERTAINTY

Where relevant, the following measurement uncertainty levels have been estimated for tests performed on the apparatus:

PARAMETER	UNCERTAINTY
Power Line Conducted Emission	+/- 2.3 dB
Radiated Emission	+/- 3.4 dB

Uncertainty figures are valid to a confidence level of 95%.

DATE: MAY 12, 2009 IC: 4168A-WA11ABG09

5. DYNAMIC FREQUENCY SELECTION

5.1. OVERVIEW

5.1.1. LIMITS

INDUSTRY CANADA

IC RSS-210 is closely harmonized with FCC Part 15 DFS rules. The deviations are as follows:

RSS-210 Issue 7 A9.4 (b) (ii) Channel Availability Check Time: ...

Additional requirements for the band 5600-5650 MHz: Until further notice, devices subject to this Section shall not be capable of transmitting in the band 5600-5650 MHz, so that Environment Canada weather radars operating in this band are protected.

RSS-210 Issue 7 A9.4 (b) (iv) **Channel closing time:** the maximum channel closing time is 260 ms.

FCC

§15.407 (h) and FCC 06-96 APPENDIX "COMPLIANCE MEASUREMENT PROCEDURES FOR UNLICENSED-NATIONAL INFORMATION INFRASTRUCTURE DEVCIES OPERATING IN THE 5250-5350 MHz AND 5470-5725 MHz BANDS INCORPORATING DYNAMIC FREQUENCY SELECTION".

Table 1: Applicability of DFS requirements prior to use of a channel

Requirement	Operatio	Operational Mode			
	Master	Client (without radar detection)	Client (with radar detection)		
Non-Occupancy Period	Yes	Not required	Yes		
DFS Detection Threshold	Yes	Not required	Yes		
Channel Availability Check Time	Yes	Not required	Not required		
Uniform Spreading	Yes	Not required	Not required		

Table 2: Applicability of DFS requirements during normal operation

Requirement	Operational I	Operational Mode			
	Master	Client	Client		
		(without DFS)	(with DFS)		
DFS Detection Threshold	Yes	Not required	Yes		
Channel Closing Transmission Time	Yes	Yes	Yes		
Channel Move Time	Yes	Yes	Yes		

Table 3: Interference Threshold values, Master or Client incorporating In-Service Monitoring

Widilitaring	
Maximum Transmit Power	Value
	(see note)
≥ 200 milliwatt	-64 dBm
< 200 milliwatt	-62 dBm

Note 1: This is the level at the input of the receiver assuming a 0 dBi receive antenna Note 2: Throughout these test procedures an additional 1 dB has been added to the amplitude of the test transmission waveforms to account for variations in measurement equipment. This will ensure that the test signal is at or above the detection threshold level to trigger a DFS response.

DATE: MAY 12, 2009 IC: 4168A-WA11ABG09

Table 4: DFS Response requirement values

Parameter	Value
Non-occupancy period	30 minutes
Channel Availability Check Time	60 seconds
Channel Move Time	10 seconds
Channel Closing Transmission Time	200 milliseconds +
	approx. 60 milliseconds
	over remaining 10 second
	period

The instant that the Channel Move Time and the Channel Closing Transmission Time begins is as follows:

For the Short pulse radar Test Signals this instant is the end of the Burst.

For the Frequency Hopping radar Test Signal, this instant is the end of the last radar burst

For the Long Pulse radar Test Signal this instant is the end of the 12 second period defining the radar transmission.

The Channel Closing Transmission Time is comprised of 200 milliseconds starting at the beginning of the Channel Move Time plus any additional intermittent control signals required to facilitate channel changes (an aggregate of approximately 60 milliseconds) during the remainder of the 10 second period. The aggregate duration of control signals will not count quiet periods in between transmissions.

Table 5 - Short Pulse Radar Test Waveforms

Table 6 Gilotti dico radai roct travolorillo								
Radar	Pulse Width	PRI	Pulses	Minimum	Minimum			
Type	(Microseconds)	(Microseconds)		Percentage of	Trials			
				Successful				
				Detection				
1	1	1428	18	60%	30			
2	1-5	150-230	23-29	60%	30			
3	6-10	200-500	16-18	60%	30			
4	11-20	200-500	12-16	60%	30			
Aggregate (F	Aggregate (Radar Types 1-4) 80% 120							

Table 6 - Long Pulse Radar Test Signal


Radar Waveform	Bursts	Pulses per Burst	Pulse Width (µsec)	Chirp Width (MHz)	PRI (µsec)	Minimum Percentage of Successful Detection	Minimum Trials
5	8-20	1-3	50-100	5-20	1000- 2000	80%	30

Table 7 – Frequency Hopping Radar Test Signal

		<u>, , , , , , , , , , , , , , , , , , , </u>			3		
Radar	Pulse	PRI	Burst	Pulses	Hopping	Minimum	Minimum
Waveform	Width	(µsec)	Length	per	Rate	Percentage of	Trials
	(µsec)		(ms)	Нор	(kHz)	Successful	
						Detection	
6	1	333	300	9	.333	70%	30

5.1.2. TEST AND MEASUREMENT SYSTEM

RADIATED METHOD SYSTEM BLOCK DIAGRAM



FCC ID: PGUWA11ABG09 IC: 4168A-WA11ABG09

DATE: MAY 12, 2009

SYSTEM OVERVIEW

The short pulse and long pulse signal generating system utilizes the NTIA software. The Vector Signal Generator has been validated by the NTIA. The hopping signal generating system utilizes the CCS simulated hopping method and system, which has been validated by the DoD, FCC and NTIA. The software selects waveform parameters from within the bounds of the signal type on a random basis using uniform distribution.

The short pulse types 2, 3 and 4, and the long pulse type 5 parameters are randomized at runtime.

The hopping type 6 pulse parameters are fixed while the hopping sequence is based on the August 2005 NTIA Hopping Frequency List. The initial starting point randomized at run-time and each subsequent starting point is incremented by 475. Each frequency in the 100-length segment is compared to the boundaries of the EUT Detection Bandwidth and the software creates a hopping burst pattern in accordance with Section 7.4.1.3 Method #2 Simulated Frequency Hopping Radar Waveform Generating Subsystem of FCC 06-96 APPENDIX. The frequency of the signal generator is incremented in 1 MHz steps from F_L to F_H for each successive trial. This incremental sequence is repeated as required to generate a minimum of 30 total trials and to maintain a uniform frequency distribution over the entire Detection Bandwidth.

The signal monitoring equipment consists of a spectrum analyzer set to display 8001 bins on the horizontal axis. The time-domain resolution is 2 msec / bin with a 16 second sweep time, meeting the 10 second short pulse reporting criteria. The aggregate ON time is calculated by multiplying the number of bins above a threshold during a particular observation period by the dwell time per bin, with the analyzer set to peak detection and max hold.

SYSTEM CALIBRATION

A 50-ohm load is connected in place of the spectrum analyzer, and the spectrum analyzer is connected to a horn antenna via a coaxial cable, with the reference level offset set to (horn antenna gain – coaxial cable loss). The signal generator is set to CW mode. The amplitude of the signal generator is adjusted to yield a level of –64 dBm as measured on the spectrum analyzer.

Without changing any of the instrument settings, the spectrum analyer is reconnected to the Common port of the Spectrum Analyzer Combiner/Divider. Measure the amplitude and calculate the difference from –64 dBm. Adjust the Reference Level Offset of the spectrum analyzer to this difference.

The spectrum analyzer displays the level of the signal generator as received at the antenna ports of the Master Device. The interference detection threshold may be varied from the calibrated value of –64 dBm and the spectrum analyzer will still indicate the level as received by the Master Device.

DATE: MAY 12, 2009 IC: 4168A-WA11ABG09

Set the signal generator to produce a radar waveform, trigger a burst manually and measure the level on the spectrum analyzer. Readjust the amplitude of the signal generator as required so that the peak level of the waveform is at a displayed level equal to the required or desired interference detection threshold. Separate signal generator amplitude settings are determined as required for each radar type.

ADJUSTMENT OF DISPLAYED TRAFFIC LEVEL

Establish a link between the Master and Slave, adjusting the distance between the units as needed to provide a suitable received level at the Master and Slave devices. Stream the video test file to generate WLAN traffic. Confirm that the WLAN traffic level, as displayed on the spectrum analyzer, is at lower amplitude than the radar detection threshold. For Master Device testing confirm that the displayed traffic does not include Slave Device traffic. For Slave Device testing confirm that the displayed traffic does not include Master Device traffic.

If a different setting of the Step Attenuators are required to meet the above conditions, perform a new System Calibration for the new Step Attenuator settings.


TEST AND MEASUREMENT EQUIPMENT

The following test and measurement equipment was utilized for the DFS tests documented in this report:

TEST EQUIPMENT LIST							
Description Manufacturer Model Serial Number Cal Due							
Spectrum Analyzer, 44 GHz	Agilent / HP	E4446A	C00986	02/03/10			
Vector signal generator, 20GHz	Agilent / HP	E8267C	C01066	11/16/09			

5.1.3. SETUP OF EUT

RADIATED METHOD EUT TEST SETUP



SUPPORT EQUIPMENT

The following test and measurement equipment was utilized for the DFS tests documented in this report:

PERIPHERAL SUPPORT EQUIPMENT LIST								
Description	Manufacturer	Model	Serial Number	FCC ID				
Notebook PC (Master)	Dell	PP18L	10657517255	DoC				
AC Adapter (Master	Lite On Technology	LA65SN0-00	CN-ODF263-	DoC				
Notebook)	Corp.		71615-687-49E					
802.11 a/b/g Wireless	Cisco	AIR-AP1231G-A-	FTX083J01D	LDK 102049				
Access Point (Master)		K9						
AC Adapter (Wireless AP)	Phihong	PSA18U-480C	P60220528A4	DoC				
802.11 A/B/G Antenna	Welch Allyn	031-0168-00	R1171	N/A				
Assembly (Slave)								
TCC Test Fixture (Slave)	Welch Allyn	851-0034-01	28	N/A				
AC Adapter (Test Fixture)	CUI, Inc.	EPAS-101W-05	02288	DoC				
Notebook PC (Slave)	Dell	PP18L	36778905757	DoC				
AC Adapter (Slave PC)	Lite On Technology	LA65SN0-00	CN-ODF263-	DoC				
	Corp.		71615-6AU-					

5.1.4. DESCRIPTION OF EUT

The EUT operates over the 5250-5350 MHz range.

The EUT is a Slave Device without Radar Detection.

The highest power level is 21 dBm EIRP.

The only antenna assembly utilized with the EUT has a max gain is 5 dBi

In lieu of a streaming full motion MP3 video file, the FCC has allowed Welch Allyn to generate WLAN traffic by downloading a file from the Master to the Slave. The file must be of sufficient length to run all required DFS testing. The file chosen was the video file TestFile.mp2 "6 $\frac{1}{2}$ Magic Hours".

DATE: MAY 12, 2009

IC: 4168A-WA11ABG09

TPC is not required since the maximum EIRP is less than 500 mW (27 dBm).

The EUT utilizes the 802.11a architecture. One nominal channel bandwidth, 20 MHz, is implemented.

The software installed in the EUT is 4.1.2 Kernel.

OVERVIEW OF MASTER DEVICE WITH RESPECT TO §15.407 (h) REQUIREMENTS

The Master Device is a Cisco Access Point, FCC ID: LDK102049. The minimum antenna gain for the Master Device is 3.5 dBi.

The rated output power of the Master unit is > 23dBm (EIRP). Therefore the required interference threshold level is –64 dBm. After correction for antenna gain and procedural adjustments, the required conducted threshold at the antenna port is –64 + 1 = -63 dBm.

The calibrated radiated DFS Detection Threshold level is set to –64 dBm. The tested level is lower than the required level hence it provides margin to the limit.


5.2. RESULTS FOR 20 MHz BANDWIDTH

5.2.1. TEST CHANNEL


All tests were performed at a channel center frequency of 5300 MHz. Measurements were performed using conducted test methods.

5.2.2. PLOTS OF RADAR WAVEFORM AND WLAN TRAFFIC

PLOTS OF RADAR WAVEFORM



PLOT OF WLAN TRAFFIC



5.2.3. MOVE AND CLOSING TIME

REPORTING NOTES

The reference marker is set at the end of last radar pulse.

The delta marker is set at the end of the last WLAN transmission following the radar pulse. This delta is the channel move time.

The aggregate channel closing transmission time is calculated as follows:

Aggregate Transmission Time = (Number of analyzer bins showing transmission) * (dwell time per bin)

The observation period over which the FCC aggregate time is calculated begins at (Reference Marker + 200 msec) and ends no earlier than (Reference Marker + 10 sec).


The observation period over which the IC aggregate time is calculated begins at (Reference Marker) and ends no earlier than (Reference Marker + 10 sec).

RESULTS


Agency	Channel Move Time	Limit
	(sec)	(sec)
FCC / IC	0.586	10

Agency	Aggregate Channel Closing Transmission Time	Limit
	(msec)	(msec)
FCC	6.0	60
IC	46.0	260

MOVE TIME




CHANNEL CLOSING TIME




AGGREGATE CHANNEL CLOSING TRANSMISSION TIME

Only intermittent transmissions are observed during the FCC aggregate monitoring period.




Only intermittent transmissions are observed during the IC aggregate monitoring period.



5.2.4. NON-OCCUPANCY


TEST RESULTS


No EUT transmissions were observed on the test channel during the 30-minute observation time.




6. SETUP PHOTOS

DYNAMIC FREQUENCY SELECTION MEASUREMENT SETUP







END OF REPORT