

Test Of: Sony Ericsson Mobile Communications AB.  
GC89 Mobile Station.  
To: OET Bulletin 65 Supplement C: (2001-01)

---

### **Appendix 3. Test Configuration Photograph**

This appendix contains the following photograph(s):

| <b>Photograph Reference Number</b> | <b>Title</b>   |
|------------------------------------|--|
| PHT/SAR_Configuration              | Test configuration for the measurement of Specific Absorption Rate (SAR) |

Test Of: Sony Ericsson Mobile Communications AB.  
GC89 Mobile Station.  
To: OET Bulletin 65 Supplement C: (2001-01)

---

**PHT/SAR\_Configuration**



Test Of: Sony Ericsson Mobile Communications AB.  
GC89 Mobile Station.  
To: OET Bulletin 65 Supplement C: (2001-01)

---

#### **Appendix 4. Calibration Data**

This appendix contains the calibration data and certificates.

| <b>Asset Number</b> | <b>Date</b>   | <b>Title</b>      |
|---------------------|---------------|-------------------|
| A1185               | 15 July 2005  | ET3DV6 – SN: 1528 |
| A1235               | 13 May 2003   | D900V2 – SN: 124  |
| A1190               | 15 April 2004 | D1800V2 – SN: 264 |
| A1186               | 10 June 2004  | ET3DV6 – SN: 1529 |

**RFI GLOBAL SERVICES LTD**

**TEST REPORT**  
**S.No. RFI/SARE1/RP70868JD08A**  
**Page 41 of 60**  
**Issue Date: 03 May 2005**

**Test Of: Sony Ericsson Mobile Communications AB.**  
**GC89 Mobile Station.**  
**To: OET Bulletin 65 Supplement C: (2001-01)**

---

This page has been left intentionally blank.

Calibration Laboratory of  
Schmid & Partner  
Engineering AG  
Zeughausstrasse 43, 8004 Zurich, Switzerland

*[Signature]*  
19/07/04

Client

RFI

**CALIBRATION CERTIFICATE**

Object(s) **ET3DV6 - SN:1528**

Calibration procedure(s) **QA CAL-01.v2  
Calibration procedure for dosimetric E-field probes**

Calibration date: **July 15, 2004**

Condition of the calibrated item **In Tolerance (according to the specific calibration document)**

This calibration certificate documents the traceability to national standards, which realize the physical units of measurements (SI).  
The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.

All calibrations have been conducted in the closed laboratory facility: environment temperature 22 +/- 2 degrees Celsius and humidity < 75%.

Calibration Equipment used (M&TE critical for calibration)

| Model Type                        | ID #           | Cal Date (Calibrated by, Certificate No.) | Scheduled Calibration  |
|-----------------------------------|----------------|---|------------------------|
| Power meter EPM E4419B            | GB41293874     | 5-May-04 (METAS, No 251-00388)            | May-05                 |
| Power sensor E4412A               | MY41495277     | 5-May-04 (METAS, No 251-00388)            | May-05                 |
| Reference 20 dB Attenuator        | SN: 5086 (20b) | 3-May-04 (METAS, No 251-00389)            | May-05                 |
| Fluke Process Calibrator Type 702 | SN: 6295803    | 8-Sep-03 (Sintrel SCS No. E-030020)       | Sep-04                 |
| Power sensor HP 8481A             | MY41092180     | 18-Sep-02 (SPEAG, in house check Oct-03)  | In house check: Oct 05 |
| RF generator HP 8684C             | US3642U01700   | 4-Aug-99 (SPEAG, in house check Aug-02)   | In house check: Aug-05 |
| Network Analyzer HP 8753E         | US37390585     | 18-Oct-01 (SPEAG, in house check Oct-03)  | In house check: Oct 05 |

|                | Name          | Function            | Signature          |
|----------------|---------------|---------------------|--------------------|
| Calibrated by: | Nico Vetterli | Technician          | <i>[Signature]</i> |
| Approved by:   | Katja Pokovic | Laboratory Director | <i>[Signature]</i> |

Date issued: July 15, 2004

This calibration certificate is issued as an intermediate solution until the accreditation process (based on ISO/IEC 17025 International Standard) for Calibration Laboratory of Schmid & Partner Engineering AG is completed.

# Probe ET3DV6

**SN:1528**

|                  |                |
|------------------|----------------|
| Manufactured:    | March 21, 2000 |
| Last calibrated: | July 29, 2003  |
| Recalibrated:    | July 15, 2004  |

**Calibrated for DASY Systems**

(Note: non-compatible with DASY2 system!)

**DASY - Parameters of Probe: ET3DV6 SN:1528****Sensitivity in Free Space**

|       |  |
|-------|--|
| NormX | 1.55 $\mu\text{V}/(\text{V}/\text{m})^2$ |
| NormY | 1.33 $\mu\text{V}/(\text{V}/\text{m})^2$ |
| NormZ | 1.40 $\mu\text{V}/(\text{V}/\text{m})^2$ |

**Diode Compression<sup>A</sup>**

|       |     |    |
|-------|-----|----|
| DCP X | 100 | mV |
| DCP Y | 100 | mV |
| DCP Z | 100 | mV |

**Sensitivity in Tissue Simulating Liquid (Conversion Factors)**

Please see Page 7.

**Boundary Effect**

**Head**                      **900 MHz**      **Typical SAR gradient: 5 % per mm**

|   |                              |        |        |
|---|------------------------------|--------|--------|
| Sensor Center to Phantom Surface Distance |                              | 3.7 mm | 4.7 mm |
| SAR <sub>be</sub> [%]                     | Without Correction Algorithm | 10.1   | 5.5    |
| SAR <sub>be</sub> [%]                     | With Correction Algorithm    | 0.0    | 0.1    |

**Head**                      **1750 MHz**      **Typical SAR gradient: 10 % per mm**

|   |                              |        |        |
|---|------------------------------|--------|--------|
| Sensor Center to Phantom Surface Distance |                              | 3.7 mm | 4.7 mm |
| SAR <sub>be</sub> [%]                     | Without Correction Algorithm | 12.9   | 8.9    |
| SAR <sub>be</sub> [%]                     | With Correction Algorithm    | 0.2    | 0.4    |

**Sensor Offset**


|                            |              |
|----------------------------|--------------|
| Probe Tip to Sensor Center | 2.7 mm       |
| Optical Surface Detection  | in tolerance |

**The reported uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k=2, which for a normal distribution corresponds to a coverage probability of approximately 95%.**




<sup>A</sup> numerical linearization parameter: uncertainty not required

# Frequency Response of E-Field

( TEM-Cell:ifi110, Waveguide R22)





### Receiving Pattern ( $\phi$ ), $\theta = 0^\circ$





**Axial Isotropy Error  $< \pm 0.2$  dB**

### Dynamic Range f(SAR<sub>head</sub>) ( Waveguide R22 )



Probe Linearity Error  $< \pm 0.2$  dB

## Conversion Factor Assessment




| f [MHz] | Validity [MHz] <sup>B</sup> | Tissue | Permittivity | Conductivity | Alpha | Depth | ConvF Uncertainty |
|---------|-----------------------------|--------|--------------|--------------|-------|-------|-------------------|
| 835     | 785-885                     | Head   | 41.5 ± 5%    | 0.90 ± 5%    | 0.65  | 1.84  | 6.23 ± 9.7% (k=2) |
| 900     | 850-950                     | Head   | 41.5 ± 5%    | 0.97 ± 5%    | 0.68  | 1.84  | 6.01 ± 9.7% (k=2) |
| 1750    | 1700-1800                   | Head   | 40.0 ± 5%    | 1.40 ± 5%    | 0.51  | 2.49  | 4.93 ± 9.7% (k=2) |
| 1900    | 1850-1950                   | Head   | 40.0 ± 5%    | 1.40 ± 5%    | 0.55  | 2.49  | 4.78 ± 9.7% (k=2) |
| 2450    | 2400-2500                   | Head   | 39.2 ± 5%    | 1.80 ± 5%    | 1.10  | 1.83  | 4.35 ± 9.7% (k=2) |
| 835     | 785-885                     | Body   | 55.2 ± 5%    | 0.97 ± 5%    | 0.57  | 2.11  | 6.07 ± 9.7% (k=2) |
| 900     | 850-950                     | Body   | 55.0 ± 5%    | 1.05 ± 5%    | 0.60  | 2.03  | 5.86 ± 9.7% (k=2) |
| 1750    | 1700-1800                   | Body   | 53.3 ± 5%    | 1.52 ± 5%    | 0.56  | 2.84  | 4.46 ± 9.7% (k=2) |
| 1900    | 1850-1950                   | Body   | 53.3 ± 5%    | 1.52 ± 5%    | 0.58  | 2.94  | 4.24 ± 9.7% (k=2) |
| 2450    | 2400-2500                   | Body   | 52.7 ± 5%    | 1.95 ± 5%    | 1.41  | 1.47  | 4.13 ± 9.7% (k=2) |

<sup>B</sup> The total standard uncertainty is calculated as root-sum-square of standard uncertainty of the Conversion Factor at calibration frequency and the standard uncertainty for the indicated frequency band.

# Deviation from Isotropy in HSL

Error ( $\theta, \phi$ ),  $f = 900$  MHz



**Spherical Isotropy Error  $< \pm 0.4$  dB**

Calibration Laboratory of  
Schmid & Partner  
Engineering AG  
Zeughausstrasse 43, 8004 Zurich, Switzerland

*Y.H. M. Adriano*  
*Curcuro*  
*23/05/03*

Client

RFI

## CALIBRATION CERTIFICATE

Object(s) **D900V2 - SN: 124**

Calibration procedure(s) **QA CAL-05.v2  
Calibration procedure for dipole validation kits**

Calibration date: **May 13, 2003**

Condition of the calibrated item **In Tolerance (according to the specific calibration document)**

This calibration statement documents traceability of M&TE used in the calibration procedures and conformity of the procedures with the ISO/IEC 17025 international standard.

All calibrations have been conducted in the closed laboratory facility: environment temperature 22 +/- 2 degrees Celsius and humidity < 75%.

Calibration Equipment used (M&TE critical for calibration)

| Model Type                | ID #       | Cal Date (Calibrated by, Certificate No.) | Scheduled Calibration  |
|---------------------------|------------|---|------------------------|
| RF generator R&S SML-03   | 100698     | 27-Mar-2002 (R&S, No. 20-92389)           | In house check: Mar-05 |
| Power sensor HP 8481A     | MY41092317 | 18-Oct-02 (Agilent, No. 20021018)         | Oct-04                 |
| Power sensor HP 8481A     | US37292783 | 30-Oct-02 (METAS, No. 252-0236)           | Oct-03                 |
| Power meter EPM E442      | GB37480704 | 30-Oct-02 (METAS, No. 252-0236)           | Oct-03                 |
| Network Analyzer HP 8753E | US38432426 | 3-May-00 (Agilent, No. 8702K064602)       | In house check: May 03 |

|                | Name           | Function            | Signature             |
|----------------|----------------|---------------------|-----------------------|
| Calibrated by: | Judith Mueller | Technician          | <i>Judith Mueller</i> |
| Approved by:   | Katja Pokovic  | Laboratory Director | <i>Katja Pokovic</i>  |

Date issued: May 13, 2003

This calibration certificate is issued as an intermediate solution until the accreditation process (based on ISO/IEC 17025 International Standard) for Calibration Laboratory of Schmid & Partner Engineering AG is completed.

# DASY

## Dipole Validation Kit

Type: D900V2

Serial: 124

Manufactured: July 4, 2001

Calibrated: May 13, 2003

## 1. Measurement Conditions

The measurements were performed in the flat section of the SAM twin phantom filled with head simulating solution of the following electrical parameters at 900 MHz:

|                        |                   |           |
|------------------------|-------------------|-----------|
| Relative Dielectricity | <b>42.1</b>       | $\pm 5\%$ |
| Conductivity           | <b>0.95 mho/m</b> | $\pm 5\%$ |

The DASY4 System with a dosimetric E-field probe ET3DV6 (SN:1507, Conversion factor 6.6 at 900 MHz) was used for the measurements.

The dipole was mounted on the small tripod so that the dipole feedpoint was positioned below the center marking of the flat phantom section and the dipole was oriented parallel to the body axis (the long side of the phantom). The standard measuring distance was 15mm from dipole center to the solution surface. The included distance holder was used during measurements for accurate distance positioning.

The coarse grid with a grid spacing of 15mm was aligned with the dipole. The 7x7x7 fine cube was chosen for cube integration.

The dipole input power (forward power) was  $250\text{mW} \pm 3\%$ . The results are normalized to 1W input power.

## 2. SAR Measurement with DASY4 System

Standard SAR-measurements were performed according to the measurement conditions described in section 1. The results (see figure supplied) have been normalized to a dipole input power of 1W (forward power). The resulting averaged SAR-values measured with the dosimetric probe ET3DV6 SN:1507 and applying the advanced extrapolation are:

|  |  |
|--|--|
| averaged over $1\text{ cm}^3$ (1 g) of tissue:   | <b>10.6 mW/g <math>\pm 16.8\%</math> (k=2)<sup>1</sup></b> |
| averaged over $10\text{ cm}^3$ (10 g) of tissue: | <b>6.76 mW/g <math>\pm 16.2\%</math> (k=2)<sup>1</sup></b> |

---

<sup>1</sup> validation uncertainty

### 3. Dipole Impedance and Return Loss

The impedance was measured at the SMA-connector with a network analyzer and numerically transformed to the dipole feedpoint. The transformation parameters from the SMA-connector to the dipole feedpoint are:

|                      |                 |                                       |
|----------------------|-----------------|---------------------------------------|
| Electrical delay:    | <b>1.381 ns</b> | (one direction)                       |
| Transmission factor: | <b>0.989</b>    | (voltage transmission, one direction) |

The dipole was positioned at the flat phantom sections according to section 1 and the distance holder was in place during impedance measurements.

|                                 |                                |
|---------------------------------|--------------------------------|
| Feedpoint impedance at 900 MHz: | $\text{Re}\{Z\} = 50.3 \Omega$ |
|---------------------------------|--------------------------------|

|  |                                |
|--|--------------------------------|
|  | $\text{Im}\{Z\} = -6.4 \Omega$ |
|--|--------------------------------|

|                        |                 |
|------------------------|-----------------|
| Return Loss at 900 MHz | <b>-24.0 dB</b> |
|------------------------|-----------------|

### 4. Measurement Conditions

The measurements were performed in the flat section of the SAM twin phantom filled with body simulating solution of the following electrical parameters at 900 MHz:

|                        |                   |           |
|------------------------|-------------------|-----------|
| Relative Dielectricity | <b>53.5</b>       | $\pm 5\%$ |
| Conductivity           | <b>1.03 mho/m</b> | $\pm 5\%$ |

The DASY4 System with a dosimetric E-field probe ET3DV6 (SN:1507, Conversion factor 6.3 at 900 MHz) was used for the measurements.

The dipole was mounted on the small tripod so that the dipole feedpoint was positioned below the center marking of the flat phantom section and the dipole was oriented parallel to the body axis (the long side of the phantom). The standard measuring distance was 15mm from dipole center to the solution surface. The included distance holder was used during measurements for accurate distance positioning.

The coarse grid with a grid spacing of 15mm was aligned with the dipole. The 7x7x7 fine cube was chosen for cube integration.

The dipole input power (forward power) was 250mW  $\pm 3\%$ . The results are normalized to 1W input power.

## 5. SAR Measurement with DASY4 System

Standard SAR-measurements were performed according to the measurement conditions described in section 4. The results (see figure supplied) have been normalized to a dipole input power of 1W (forward power). The resulting averaged SAR-values measured with the dosimetric probe ET3DV6 SN:1507 and applying the advanced extrapolation are:

|  |   |
|--|---|
| averaged over 1 cm <sup>3</sup> (1 g) of tissue:   | <b>11.0 mW/g ± 16.8 % (k=2)<sup>2</sup></b> |
| averaged over 10 cm <sup>3</sup> (10 g) of tissue: | <b>7.12 mW/g ± 16.2 % (k=2)<sup>2</sup></b> |

## 6. Dipole Impedance and Return Loss

The dipole was positioned at the flat phantom sections according to section 4 and the distance holder was in place during impedance measurements.

|                                 |                        |
|---------------------------------|------------------------|
| Feedpoint impedance at 900 MHz: | <b>Re{Z} = 46.2 Ω</b>  |
|                                 | <b>Im {Z} = -8.2 Ω</b> |
| Return Loss at 900 MHz          | <b>-20.6 dB</b>        |

## 7. Handling

Do not apply excessive force to the dipole arms, because they might bend. Bending of the dipole arms stresses the soldered connections near the feedpoint leading to a damage of the dipole.

## 8. Design

The dipole is made of standard semirigid coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC-signals.

## 9. Power Test

After long term use with 100W radiated power, only a slight warming of the dipole near the feedpoint can be measured.

---

<sup>2</sup> validation uncertainty

Date/Time: 05/09/03 15:50:49

Test Laboratory: SPEAG, Zurich, Switzerland  
 File Name: SN0124\_SN1507\_HSL900\_090503da4.da4

**DUT: Dipole 900 MHz; Type: D900V2; Serial: D900V2 - SN124**  
**Program: Dipole Calibration**


Communication System: CW-900; Frequency: 900 MHz; Duty Cycle: 1:1  
 Medium: HSL 900 MHz ( $\sigma = 0.95$  mho/m,  $\epsilon_r = 42.07$ ,  $\rho = 1000$  kg/m<sup>3</sup>)  
 Phantom section: Flat Section  
 Measurement Standard: DASYS4 (High Precision Assessment)

DASY4 Configuration:

- Probe: ET3DV6 - SN1507; ConvF(6.6, 6.6, 6.6); Calibrated: 1/18/2003
- Sensor-Surface: 4mm (Mechanical Surface Detection)
- Electronics: DAE3 - SN411; Calibrated: 1/16/2003
- Phantom: SAM with CRP - TP1006; Type: SAM 4.0; Serial: TP:1006
- Measurement SW: DASYS4, V4.1 Build 47; Postprocessing SW: SEMCAD, V1.6 Build 115

**Pin = 250 mW; d = 15 mm/Area Scan (81x81x1):** Measurement grid: dx=15mm, dy=15mm  
 Reference Value = 57.1 V/m  
 Power Drift = 0.02 dB  
 Maximum value of SAR = 2.82 mW/g

**Pin = 250 mW; d = 15 mm/Zoom Scan (7x7x7)/Cube 0:** Measurement grid: dx=5mm, dy=5mm, dz=5mm  
 Peak SAR (extrapolated) = 3.88 W/kg  
 SAR(1 g) = 2.64 mW/g; SAR(10 g) = 1.69 mW/g  
 Reference Value = 57.1 V/m  
 Power Drift = 0.02 dB  
 Maximum value of SAR = 2.83 mW/g



0 dB = 2.83mW/g

9 May 2003 11:16:17

CHI S11 1 U FS

1: 50.324  $\Omega$  -6.3711  $\Omega$  27.756 pF


900.000 000 MHz

Del

Cor

Avg  
16


↑



CH2 S11 LOG 4 dB/REF -20 dB 1:-23.950 dB 900.000 000 MHz

Cor

↑



CENTER 900.000 000 MHz

SPAN 400.000 000 MHz

Date/Time: 05/13/03 11:27:28

Test Laboratory: SPEAG, Zurich, Switzerland  
 File Name: SN124\_SN1507\_M900\_130503.da4

**DUT: Dipole 900 MHz; Type: D900V2; Serial: D900V2 - SN124**  
**Program: Dipole Calibration**


Communication System: CW-900; Frequency: 900 MHz; Duty Cycle: 1:1  
 Medium: Muscle 900 MHz ( $\sigma = 1.03$  mho/m,  $\epsilon_r = 53.48$ ,  $\rho = 1000$  kg/m<sup>3</sup>)  
 Phantom section: Flat Section  
 Measurement Standard: DASy4 (High Precision Assessment)

DASy4 Configuration:

- Probe: ET3DV6 - SN1507; ConvF(6.3, 6.3, 6.3); Calibrated: 1/18/2003
- Sensor-Surface: 4mm (Mechanical Surface Detection)
- Electronics: DAE3 - SN411; Calibrated: 1/16/2003
- Phantom: SAM with CRP - TP1006; Type: SAM 4.0; Serial: TP:1006
- Measurement SW: DASy4, V4.1 Build 47; Postprocessing SW: SEMCAD, V1.6 Build 115

**Pin = 250 mW; d = 15 mm/Area Scan (81x81x1):** Measurement grid: dx=15mm, dy=15mm  
 Reference Value = 56 V/m  
 Power Drift = 0.007 dB  
 Maximum value of SAR = 2.94 mW/g

**Pin = 250 mW; d = 15 mm/Zoom Scan (7x7x7)/Cube 0:** Measurement grid: dx=5mm, dy=5mm, dz=5mm  
 Peak SAR (extrapolated) = 3.97 W/kg  
 SAR(1 g) = 2.75 mW/g; SAR(10 g) = 1.78 mW/g  
 Reference Value = 56 V/m  
 Power Drift = 0.007 dB  
 Maximum value of SAR = 2.97 mW/g



0 dB = 2.97mW/g

124  
Body


13 May 2003 09:21:44

CH1 S11 1 U FS

1: 46.223  $\Omega$  -8.1541  $\Omega$  21.661 pF


900.000 000 MHz

Del  
PRM  
Cor  
Avg  
16  
↑



CH2 S11 LOG 4 dB/REF -20 dB 1:-20.546 dB 900.000 000 MHz

PRM  
Cor  
↑



CENTER 900.000 000 MHz

SPAN 400.000 000 MHz

Calibration Laboratory of  
Schmid & Partner  
Engineering AG  
Zeughausstrasse 43, 8004 Zurich, Switzerland

*Handwritten signature and date:*  
07/09/04

Client

RFI

**CALIBRATION CERTIFICATE**

Object(s) **D1800V2 - SN:264**

Calibration procedure(s) **QA CAL-05.v2  
Calibration procedure for dipole validation kits**

Calibration date: **April 15, 2004**

Condition of the calibrated item **In Tolerance (according to the specific calibration document)**

This calibration statement documents traceability of M&TE used in the calibration procedures and conformity of the procedures with the ISO/IEC 17025 international standard.

All calibrations have been conducted in the closed laboratory facility: environment temperature 22 +/- 2 degrees Celsius and humidity < 75%.

Calibration Equipment used (M&TE critical for calibration)

| Model Type                | ID #       | Cal Date (Calibrated by, Certificate No.) | Scheduled Calibration  |
|---------------------------|------------|---|------------------------|
| Power meter EPM E442      | GB37480704 | 6-Nov-03 (METAS, No. 252-0254)            | Nov-04                 |
| Power sensor HP 8481A     | US37292783 | 6-Nov-03 (METAS, No. 252-0254)            | Nov-04                 |
| Power sensor HP 8481A     | MY41092317 | 18-Oct-02 (Agilent, No. 20021018)         | Oct-04                 |
| RF generator R&S SML-03   | 100698     | 27-Mar-2002 (R&S, No. 20-92389)           | In house check: Mar-05 |
| Network Analyzer HP 8753E | US37390585 | 18-Oct-01 (SPEAG, in house check Nov-03)  | In house check: Oct 05 |

|                | Name           | Function            | Signature          |
|----------------|----------------|---------------------|--------------------|
| Calibrated by: | Judith Mueller | Technician          | <i>[Signature]</i> |
| Approved by:   | Katja Pokovic  | Laboratory Director | <i>[Signature]</i> |

Date issued: April 21, 2004

This calibration certificate is issued as an intermediate solution until the accreditation process (based on ISO/IEC 17025 International Standard) for Calibration Laboratory of Schmid & Partner Engineering AG is completed.

# DASY

## Dipole Validation Kit

Type: D1800V2

Serial: 264

Manufactured: March 5, 2000

Calibrated: April 15, 2004

## 1. Measurement Conditions

The measurements were performed in the flat section of the SAM twin phantom filled with **head simulating solution** of the following electrical parameters at 1800 MHz:

|                        |                   |           |
|------------------------|-------------------|-----------|
| Relative Dielectricity | <b>40.6</b>       | $\pm 5\%$ |
| Conductivity           | <b>1.36 mho/m</b> | $\pm 5\%$ |

The DASY4 System with a dosimetric E-field probe ET3DV6 (SN:1507, Conversion factor 5.08 at 1800 MHz) was used for the measurements.

The dipole was mounted on the small tripod so that the dipole feedpoint was positioned below the center marking of the flat phantom section and the dipole was oriented parallel to the body axis (the long side of the phantom). The standard measuring distance was 10mm from dipole center to the solution surface. The included distance spacer was used during measurements for accurate distance positioning.

The coarse grid with a grid spacing of 15mm was aligned with the dipole. The 7x7x7 fine cube was chosen for cube integration.

The dipole input power (forward power) was  $250 \text{ mW} \pm 3\%$ . The results are normalized to 1W input power.

## 2. SAR Measurement with DASY4 System

Standard SAR-measurements were performed according to the measurement conditions described in section 1. The results (see figure supplied) have been normalized to a dipole input power of 1W (forward power). The resulting averaged SAR-values measured with the dosimetric probe ET3DV6 SN:1507 and applying the advanced extrapolation are:

|   |  |
|---|--|
| averaged over $1 \text{ cm}^3$ (1 g) of tissue:   | <b>37.2 mW/g <math>\pm 16.8\%</math> (k=2)<sup>1</sup></b> |
| averaged over $10 \text{ cm}^3$ (10 g) of tissue: | <b>19.9 mW/g <math>\pm 16.2\%</math> (k=2)<sup>1</sup></b> |

---

<sup>1</sup> validation uncertainty



Test Laboratory: SPEAG, Zurich, Switzerland

**DUT: Dipole 1800 MHz; Type: D1800V2; Serial: D1800V2 - SN264**

Communication System: CW-1800; Frequency: 1800 MHz; Duty Cycle: 1:1

Medium: HSL 1800 MHz;

Medium parameters used:  $f = 1800$  MHz;  $\sigma = 1.36$  mho/m;  $\epsilon_r = 40.6$ ;  $\rho = 1000$  kg/m<sup>3</sup>

Phantom section: Flat Section

Measurement Standard: DASY4 (High Precision Assessment)

DASY4 Configuration:

- Probe: ET3DV6 - SN1507; ConvF(5.08, 5.08, 5.08); Calibrated: 1/23/2004
- Sensor-Surface: 4mm (Mechanical Surface Detection)
- Electronics: DAE3 Sn411; Calibrated: 11/6/2003
- Phantom: SAM with CRP - TP1006; Type: SAM 4.0; Serial: TP:1006;
- Measurement SW: DASY4, V4.2 Build 44; Postprocessing SW: SEMCAD, V1.8 Build 112

**Pin = 250 mW; d = 10 mm/Area Scan (81x81x1):** Measurement grid: dx=15mm, dy=15mm

Reference Value = 90.6 V/m; Power Drift = 0.0 dB

Maximum value of SAR (interpolated) = 10.6 mW/g


**Pin = 250 mW; d = 10 mm/Zoom Scan (7x7x7)/Cube 0:** Measurement grid: dx=5mm, dy=5mm, dz=5mm

Reference Value = 90.6 V/m; Power Drift = 0.0 dB

Maximum value of SAR (measured) = 10.5 mW/g

Peak SAR (extrapolated) = 16.3 W/kg

**SAR(1 g) = 9.31 mW/g; SAR(10 g) = 4.98 mW/g**



0 dB = 10.5mW/g

264  
Hod

15 Apr 2004 08:46:49

CH1 S11 1 U FS

1: 46.924  $\Omega$  -5.8965  $\Omega$  14.995 pF


1 800.000 000 MHz

Del

Cor

Avg  
16


↑



CH2 S11 LOG 5 dB/REF -20 dB 1:-23.289 dB 1 800.000 000 MHz

Cor

↑



CENTER 1 800.000 000 MHz

SPAN 400.000 000 MHz



Test Laboratory: SPEAG, Zurich, Switzerland

**DUT: Dipole 1800 MHz; Type: D1800V2; Serial: D1800V2 - SN264**

Communication System: CW-1800; Frequency: 1800 MHz; Duty Cycle: 1:1

Medium: Muscle 1800 MHz;

Medium parameters used:  $f = 1800$  MHz;  $\sigma = 1.49$  mho/m;  $\epsilon_r = 52.6$ ;  $\rho = 1000$  kg/m<sup>3</sup>

Phantom section: Flat Section

Measurement Standard: DASY4 (High Precision Assessment)

DASY4 Configuration:

- Probe: ET3DV6 - SN1507; ConvF(4.61, 4.61, 4.61); Calibrated: 1/23/2004
- Sensor-Surface: 4mm (Mechanical Surface Detection)
- Electronics: DAE3 Sn411; Calibrated: 11/6/2003
- Phantom: SAM with CRP - TP1006; Type: SAM 4.0; Serial: TP:1006;
- Measurement SW: DASY4, V4.2 Build 44; Postprocessing SW: SEMCAD, V1.8 Build 112

**Pin = 250 mW; d = 10 mm/Area Scan (81x81x1):** Measurement grid: dx=15mm, dy=15mm

Reference Value = 87.5 V/m; Power Drift = 0.0 dB

Maximum value of SAR (interpolated) = 10.5 mW/g


**Pin = 250 mW; d = 10 mm/Zoom Scan (7x7x7)/Cube 0:** Measurement grid: dx=5mm, dy=5mm, dz=5mm

Reference Value = 87.5 V/m; Power Drift = 0.0 dB

Maximum value of SAR (measured) = 10.5 mW/g

Peak SAR (extrapolated) = 15.4 W/kg

**SAR(1 g) = 9.25 mW/g; SAR(10 g) = 5 mW/g**



264  
Body

14 Apr 2004 10:07:36


CH1 S11 1 U FS 1: 44.342  $\Omega$  -5.7285  $\Omega$  15.435 pF 1 800.000 000 MHz

De1

Cor

Avg  
16


↑



CH2 S11 LOG 5 dB/REF -20 dB 1:-21.259 dB 1 800.000 000 MHz

Cor

↑



CENTER 1 800.000 000 MHz

SPAN 400.000 000 MHz