

|                                |                                                                 |                                            |            |
|--------------------------------|-----------------------------------------------------------------|--------------------------------------------|------------|
| <b>SONY</b>                    | Sony Mobile Communications (China) Co.,<br>Ltd. Test Laboratory | Report No.: TARC-PM-0821-BV-<br>SAR-FCC-02 |            |
| PM-0821-BV SAR FCC Test Report |                                                                 | Edition 2                                  | Revision 0 |

**Calibration Laboratory of  
Schmid & Partner  
Engineering AG**  
Zeughausstrasse 43, 8004 Zurich, Switzerland



S Schweizerischer Kalibrierdienst  
C Service suisse d'étalonnage  
S Servizio svizzero di taratura  
S Swiss Calibration Service

Accredited by the Swiss Accreditation Service (SAS)  
The Swiss Accreditation Service is one of the signatories to the EA  
Multilateral Agreement for the recognition of calibration certificates

Accreditation No.: **SCS 108**

Client **Sony Mobile Factory (PTT)**

Certificate No: **D900V2-1d065\_Feb14**

## CALIBRATION CERTIFICATE

Object **D900V2 - SN: 1d065**

Calibration procedure(s) **QA CAL-05.v9  
Calibration procedure for dipole validation kits above 700 MHz**

Calibration date: **February 21, 2014**

This calibration certificate documents the traceability to national standards, which realize the physical units of measurements (SI).  
The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.

All calibrations have been conducted in the closed laboratory facility: environment temperature  $(22 \pm 3)^\circ\text{C}$  and humidity  $< 70\%$ .

Calibration Equipment used (M&TE critical for calibration)

| Primary Standards           | ID #               | Cal Date (Certificate No.)     | Scheduled Calibration |
|-----------------------------|--------------------|--------------------------------|-----------------------|
| Power meter EPM-442A        | GB37480704         | 09-Oct-13 (No. 217-01827)      | Oct-14                |
| Power sensor HP 8481A       | US37292783         | 09-Oct-13 (No. 217-01827)      | Oct-14                |
| Power sensor HP 8481A       | MY41092317         | 09-Oct-13 (No. 217-01828)      | Oct-14                |
| Reference 20 dB Attenuator  | SN: 5058 (20k)     | 04-Apr-13 (No. 217-01736)      | Apr-14                |
| Type-N mismatch combination | SN: 5047.3 / 06327 | 04-Apr-13 (No. 217-01739)      | Apr-14                |
| Reference Probe ES3DV3      | SN: 3205           | 30-Dec-13 (No. ES3-3205_Dec13) | Dec-14                |
| DAE4                        | SN: 601            | 25-Apr-13 (No. DAE4-601_Apr13) | Apr-14                |

| Secondary Standards       | ID #             | Check Date (in house)             | Scheduled Check        |
|---------------------------|------------------|-----------------------------------|------------------------|
| RF generator R&S SMT-06   | 100005           | 04-Aug-99 (in house check Oct-13) | In house check: Oct-16 |
| Network Analyzer HP 8753E | US37390585 S4206 | 19-Oct-01 (in house check Oct-13) | In house check: Oct-14 |

|                |                              |                                          |               |
|----------------|------------------------------|------------------------------------------|---------------|
| Calibrated by: | Name<br><b>Leif Klysner</b>  | Function<br><b>Laboratory Technician</b> | Signature<br> |
| Approved by:   | Name<br><b>Katja Pokovic</b> | Function<br><b>Technical Manager</b>     | Signature<br> |

Issued: February 21, 2014

This calibration certificate shall not be reproduced except in full without written approval of the laboratory.

Certificate No: D900V2-1d065\_Feb14

Page 1 of 8

|                                |                                                                 |                                        |            |
|--------------------------------|-----------------------------------------------------------------|----------------------------------------|------------|
| <b>SONY</b>                    | Sony Mobile Communications (China) Co.,<br>Ltd. Test Laboratory | Report No.: TARC-PM-0821-BV-SAR-FCC-02 |            |
| PM-0821-BV SAR FCC Test Report |                                                                 | Edition 2                              | Revision 0 |

**Calibration Laboratory of**  
Schmid & Partner  
Engineering AG  
Zeughausstrasse 43, 8004 Zurich, Switzerland



**S** Schweizerischer Kalibrierdienst  
**C** Service suisse d'étalonnage  
**S** Servizio svizzero di taratura  
**S** Swiss Calibration Service

Accredited by the Swiss Accreditation Service (SAS)  
The Swiss Accreditation Service is one of the signatories to the EA  
Multilateral Agreement for the recognition of calibration certificates

Accreditation No.: **SCS 108**

**Glossary:**

TSL tissue simulating liquid  
ConvF sensitivity in TSL / NORM x,y,z  
N/A not applicable or not measured

**Calibration is Performed According to the Following Standards:**

- IEEE Std 1528-2013, "IEEE Recommended Practice for Determining the Peak Spatial-Averaged Specific Absorption Rate (SAR) in the Human Head from Wireless Communications Devices: Measurement Techniques", June 2013
- IEC 62209-1, "Procedure to measure the Specific Absorption Rate (SAR) for hand-held devices used in close proximity to the ear (frequency range of 300 MHz to 3 GHz)", February 2005
- KDB 865664, "SAR Measurement Requirements for 100 MHz to 6 GHz"

**Additional Documentation:**

- DASY4/5 System Handbook

**Methods Applied and Interpretation of Parameters:**

- Measurement Conditions:** Further details are available from the Validation Report at the end of the certificate. All figures stated in the certificate are valid at the frequency indicated.
- Antenna Parameters with TSL:** The dipole is mounted with the spacer to position its feed point exactly below the center marking of the flat phantom section, with the arms oriented parallel to the body axis.
- Feed Point Impedance and Return Loss:** These parameters are measured with the dipole positioned under the liquid filled phantom. The impedance stated is transformed from the measurement at the SMA connector to the feed point. The Return Loss ensures low reflected power. No uncertainty required.
- Electrical Delay:** One-way delay between the SMA connector and the antenna feed point. No uncertainty required.
- SAR measured:** SAR measured at the stated antenna input power.
- SAR normalized:** SAR as measured, normalized to an input power of 1 W at the antenna connector.
- SAR for nominal TSL parameters:** The measured TSL parameters are used to calculate the nominal SAR result.

The reported uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor  $k=2$ , which for a normal distribution corresponds to a coverage probability of approximately 95%.

|                                |                                                                 |                                        |            |
|--------------------------------|-----------------------------------------------------------------|----------------------------------------|------------|
| <b>SONY</b>                    | Sony Mobile Communications (China) Co.,<br>Ltd. Test Laboratory | Report No.: TARC-PM-0821-BV-SAR-FCC-02 |            |
| PM-0821-BV SAR FCC Test Report |                                                                 | Edition 2                              | Revision 0 |

### Measurement Conditions

DASY system configuration, as far as not given on page 1.

|                              |                        |             |
|------------------------------|------------------------|-------------|
| DASY Version                 | DASY5                  | V52.8.7     |
| Extrapolation                | Advanced Extrapolation |             |
| Phantom                      | Modular Flat Phantom   |             |
| Distance Dipole Center - TSL | 15 mm                  | with Spacer |
| Zoom Scan Resolution         | dx, dy, dz = 5 mm      |             |
| Frequency                    | 900 MHz $\pm$ 1 MHz    |             |

### Head TSL parameters

The following parameters and calculations were applied.

|                                         | Temperature         | Permittivity   | Conductivity         |
|-----------------------------------------|---------------------|----------------|----------------------|
| Nominal Head TSL parameters             | 22.0 °C             | 41.5           | 0.97 mho/m           |
| Measured Head TSL parameters            | (22.0 $\pm$ 0.2) °C | 40.4 $\pm$ 6 % | 0.95 mho/m $\pm$ 6 % |
| Head TSL temperature change during test | < 0.5 °C            | ----           | ----                 |

### SAR result with Head TSL

|                                                       |                    |                              |
|-------------------------------------------------------|--------------------|------------------------------|
| SAR averaged over 1 cm <sup>3</sup> (1 g) of Head TSL | Condition          |                              |
| SAR measured                                          | 250 mW input power | 2.59 W/kg                    |
| SAR for nominal Head TSL parameters                   | normalized to 1W   | 10.5 W/kg $\pm$ 17.0 % (k=2) |

|                                                         |                    |                              |
|---------------------------------------------------------|--------------------|------------------------------|
| SAR averaged over 10 cm <sup>3</sup> (10 g) of Head TSL | condition          |                              |
| SAR measured                                            | 250 mW input power | 1.66 W/kg                    |
| SAR for nominal Head TSL parameters                     | normalized to 1W   | 6.70 W/kg $\pm$ 16.5 % (k=2) |

### Body TSL parameters

The following parameters and calculations were applied.

|                                         | Temperature         | Permittivity   | Conductivity         |
|-----------------------------------------|---------------------|----------------|----------------------|
| Nominal Body TSL parameters             | 22.0 °C             | 55.0           | 1.05 mho/m           |
| Measured Body TSL parameters            | (22.0 $\pm$ 0.2) °C | 53.9 $\pm$ 6 % | 1.03 mho/m $\pm$ 6 % |
| Body TSL temperature change during test | < 0.5 °C            | ----           | ----                 |

### SAR result with Body TSL

|                                                       |                    |                              |
|-------------------------------------------------------|--------------------|------------------------------|
| SAR averaged over 1 cm <sup>3</sup> (1 g) of Body TSL | Condition          |                              |
| SAR measured                                          | 250 mW input power | 2.60 W/kg                    |
| SAR for nominal Body TSL parameters                   | normalized to 1W   | 10.5 W/kg $\pm$ 17.0 % (k=2) |

|                                                         |                    |                              |
|---------------------------------------------------------|--------------------|------------------------------|
| SAR averaged over 10 cm <sup>3</sup> (10 g) of Body TSL | condition          |                              |
| SAR measured                                            | 250 mW input power | 1.68 W/kg                    |
| SAR for nominal Body TSL parameters                     | normalized to 1W   | 6.78 W/kg $\pm$ 16.5 % (k=2) |

|                                |                                                                 |                                            |            |
|--------------------------------|-----------------------------------------------------------------|--------------------------------------------|------------|
| <b>SONY</b>                    | Sony Mobile Communications (China) Co.,<br>Ltd. Test Laboratory | Report No.: TARC-PM-0821-BV-<br>SAR-FCC-02 |            |
| PM-0821-BV SAR FCC Test Report |                                                                 | Edition 2                                  | Revision 0 |

## Appendix

### Antenna Parameters with Head TSL

|                                      |                                 |
|--------------------------------------|---------------------------------|
| Impedance, transformed to feed point | 51.7 $\Omega$ - 0.9 $\mu\Omega$ |
| Return Loss                          | - 34.5 dB                       |

### Antenna Parameters with Body TSL

|                                      |                                 |
|--------------------------------------|---------------------------------|
| Impedance, transformed to feed point | 46.7 $\Omega$ - 3.1 $\mu\Omega$ |
| Return Loss                          | - 26.6 dB                       |

### General Antenna Parameters and Design

|                                  |          |
|----------------------------------|----------|
| Electrical Delay (one direction) | 1.413 ns |
|----------------------------------|----------|

After long term use with 100W radiated power, only a slight warming of the dipole near the feedpoint can be measured.

The dipole is made of standard semirigid coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC-signals. On some of the dipoles, small end caps are added to the dipole arms in order to improve matching when loaded according to the position as explained in the "Measurement Conditions" paragraph. The SAR data are not affected by this change. The overall dipole length is still according to the Standard.

No excessive force must be applied to the dipole arms, because they might bend or the soldered connections near the feedpoint may be damaged.

### Additional EUT Data

|                 |                |
|-----------------|----------------|
| Manufactured by | SPEAG          |
| Manufactured on | March 08, 2007 |

|                                |                                                                 |                                            |            |
|--------------------------------|-----------------------------------------------------------------|--------------------------------------------|------------|
| <b>SONY</b>                    | Sony Mobile Communications (China) Co.,<br>Ltd. Test Laboratory | Report No.: TARC-PM-0821-BV-<br>SAR-FCC-02 |            |
| PM-0821-BV SAR FCC Test Report |                                                                 | Edition 2                                  | Revision 0 |

## DASY5 Validation Report for Head TSL

Date: 21.02.2014

Test Laboratory: SPEAG, Zurich, Switzerland

**DUT: Dipole 900 MHz; Type: D900V2; Serial: D900V2 - SN: 1d065**

Communication System: UID 0 - CW; Frequency: 900 MHz

Medium parameters used:  $f = 900 \text{ MHz}$ ;  $\sigma = 0.95 \text{ S/m}$ ;  $\epsilon_r = 40.4$ ;  $\rho = 1000 \text{ kg/m}^3$

Phantom section: Flat Section

Measurement Standard: DASY5 (IEEE/IEC/ANSI C63.19-2007)

DASY52 Configuration:

- Probe: ES3DV3 - SN3205; ConvF(6.09, 6.09, 6.09); Calibrated: 30.12.2013;
- Sensor-Surface: 3mm (Mechanical Surface Detection)
- Electronics: DAE4 Sn601; Calibrated: 25.04.2013
- Phantom: Flat Phantom 4.9L; Type: QD000P49AA; Serial: 1001
- DASY52 52.8.7(1137); SEMCAD X 14.6.10(7164)

### Dipole Calibration for Head Tissue/Pin=250 mW, d=15mm/Zoom Scan (7x7x7)/Cube 0:


Measurement grid: dx=5mm, dy=5mm, dz=5mm

Reference Value = 58.317 V/m; Power Drift = 0.00 dB

Peak SAR (extrapolated) = 3.92 W/kg


**SAR(1 g) = 2.59 W/kg; SAR(10 g) = 1.66 W/kg**

Maximum value of SAR (measured) = 3.04 W/kg



|                                |                                                                 |                                        |            |
|--------------------------------|-----------------------------------------------------------------|----------------------------------------|------------|
| <b>SONY</b>                    | Sony Mobile Communications (China) Co.,<br>Ltd. Test Laboratory | Report No.: TARC-PM-0821-BV-SAR-FCC-02 |            |
| PM-0821-BV SAR FCC Test Report |                                                                 | Edition 2                              | Revision 0 |

Impedance Measurement Plot for Head TSL



|                                |                                                                 |                                            |            |
|--------------------------------|-----------------------------------------------------------------|--------------------------------------------|------------|
| <b>SONY</b>                    | Sony Mobile Communications (China) Co.,<br>Ltd. Test Laboratory | Report No.: TARC-PM-0821-BV-<br>SAR-FCC-02 |            |
| PM-0821-BV SAR FCC Test Report |                                                                 | Edition 2                                  | Revision 0 |

## DASY5 Validation Report for Body TSL

Date: 20.02.2014

Test Laboratory: SPEAG, Zurich, Switzerland

**DUT: Dipole 900 MHz; Type: D900V2; Serial: D900V2 - SN: 1d065**

Communication System: UID 0 - CW; Frequency: 900 MHz

Medium parameters used:  $f = 900$  MHz;  $\sigma = 1.03$  S/m;  $\epsilon_r = 53.9$ ;  $\rho = 1000$  kg/m<sup>3</sup>

Phantom section: Flat Section

Measurement Standard: DASY5 (IEEE/IEC/ANSI C63.19-2007)

DASY52 Configuration:

- Probe: ES3DV3 - SN3205; ConvF(5.98, 5.98, 5.98); Calibrated: 30.12.2013;
- Sensor-Surface: 3mm (Mechanical Surface Detection)
- Electronics: DAE4 Sn601; Calibrated: 25.04.2013
- Phantom: Flat Phantom 4.9L; Type: QD000P49AA; Serial: 1001
- DASY52 52.8.7(1137); SEMCAD X 14.6.10(7164)

### Dipole Calibration for Body Tissue/Pin=250 mW, d=15mm 2/Zoom Scan (7x7x7)/Cube 0:


Measurement grid: dx=5mm, dy=5mm, dz=5mm

Reference Value = 56.065 V/m; Power Drift = 0.01 dB

Peak SAR (extrapolated) = 3.87 W/kg


**SAR(1 g) = 2.6 W/kg; SAR(10 g) = 1.68 W/kg**

Maximum value of SAR (measured) = 3.06 W/kg



|                                |                                                                 |                                        |            |
|--------------------------------|-----------------------------------------------------------------|----------------------------------------|------------|
| <b>SONY</b>                    | Sony Mobile Communications (China) Co.,<br>Ltd. Test Laboratory | Report No.: TARC-PM-0821-BV-SAR-FCC-02 |            |
| PM-0821-BV SAR FCC Test Report |                                                                 | Edition 2                              | Revision 0 |

Impedance Measurement Plot for Body TSL





|                                |                                                                 |                                            |            |
|--------------------------------|-----------------------------------------------------------------|--------------------------------------------|------------|
| <b>SONY</b>                    | Sony Mobile Communications (China) Co.,<br>Ltd. Test Laboratory | Report No.: TARC-PM-0821-BV-<br>SAR-FCC-02 |            |
| PM-0821-BV SAR FCC Test Report |                                                                 | Edition 2                                  | Revision 0 |

**Calibration Laboratory of  
Schmid & Partner  
Engineering AG**  
Zeughausstrasse 43, 8004 Zurich, Switzerland



**S** Schweizerischer Kalibrierdienst  
**C** Service suisse d'étalonnage  
**S** Servizio svizzero di taratura  
**S** Swiss Calibration Service

Accredited by the Swiss Accreditation Service (SAS)  
The Swiss Accreditation Service is one of the signatories to the EA  
Multilateral Agreement for the recognition of calibration certificates

Accreditation No.: **SCS 108**

Client **Sony Mobile Factory (PTT)**

Certificate No: **D835V2-4d061\_Feb14**

## CALIBRATION CERTIFICATE

Object **D835V2 - SN: 4d061**

Calibration procedure(s) **QA CAL-05.v9**  
**Calibration procedure for dipole validation kits above 700 MHz**

Calibration date: **February 21, 2014**

This calibration certificate documents the traceability to national standards, which realize the physical units of measurements (SI).  
The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.

All calibrations have been conducted in the closed laboratory facility: environment temperature ( $22 \pm 3$ )°C and humidity < 70%.

Calibration Equipment used (M&TE critical for calibration)

| Primary Standards           | ID #               | Cal Date (Certificate No.)     | Scheduled Calibration |
|-----------------------------|--------------------|--------------------------------|-----------------------|
| Power meter EPM-442A        | GB37480704         | 09-Oct-13 (No. 217-01827)      | Oct-14                |
| Power sensor HP 8481A       | US37292783         | 09-Oct-13 (No. 217-01827)      | Oct-14                |
| Power sensor HP 8481A       | MY41092317         | 09-Oct-13 (No. 217-01828)      | Oct-14                |
| Reference 20 dB Attenuator  | SN: 5058 (20k)     | 04-Apr-13 (No. 217-01736)      | Apr-14                |
| Type-N mismatch combination | SN: 5047.3 / 06327 | 04-Apr-13 (No. 217-01739)      | Apr-14                |
| Reference Probe ES3DV3      | SN: 3205           | 30-Dec-13 (No. ES3-3205_Dec13) | Dec-14                |
| DAE4                        | SN: 601            | 25-Apr-13 (No. DAE4-601_Apr13) | Apr-14                |

| Secondary Standards       | ID #             | Check Date (in house)             | Scheduled Check        |
|---------------------------|------------------|-----------------------------------|------------------------|
| RF generator R&S SMT-06   | 100005           | 04-Aug-99 (in house check Oct-13) | In house check: Oct-16 |
| Network Analyzer HP 8753E | US37390585 S4206 | 18-Oct-01 (in house check Oct-13) | In house check: Oct-14 |

|                |                              |                                          |               |
|----------------|------------------------------|------------------------------------------|---------------|
| Calibrated by: | Name<br><b>Leif Klysner</b>  | Function<br><b>Laboratory Technician</b> | Signature<br> |
| Approved by:   | Name<br><b>Katja Pokovic</b> | Function<br><b>Technical Manager</b>     | Signature<br> |

Issued: February 21, 2014

This calibration certificate shall not be reproduced except in full without written approval of the laboratory.

Certificate No: **D835V2-4d061\_Feb14**

Page 1 of 8

|                                |                                                                 |                                            |            |
|--------------------------------|-----------------------------------------------------------------|--------------------------------------------|------------|
| <b>SONY</b>                    | Sony Mobile Communications (China) Co.,<br>Ltd. Test Laboratory | Report No.: TARC-PM-0821-BV-<br>SAR-FCC-02 |            |
| PM-0821-BV SAR FCC Test Report |                                                                 | Edition 2                                  | Revision 0 |

**Calibration Laboratory of**  
Schmid & Partner  
Engineering AG  
Zeughausstrasse 43, 8004 Zurich, Switzerland



**S** Schweizerischer Kalibrierdienst  
**C** Service suisse d'étalonnage  
**S** Servizio svizzero di taratura  
**S** Swiss Calibration Service

Accredited by the Swiss Accreditation Service (SAS)  
The Swiss Accreditation Service is one of the signatories to the EA  
Multilateral Agreement for the recognition of calibration certificates

Accreditation No.: **SCS 108**

**Glossary:**

TSL tissue simulating liquid  
ConvF sensitivity in TSL / NORM x,y,z  
N/A not applicable or not measured

**Calibration is Performed According to the Following Standards:**

- IEEE Std 1528-2013, "IEEE Recommended Practice for Determining the Peak Spatial-Averaged Specific Absorption Rate (SAR) in the Human Head from Wireless Communications Devices: Measurement Techniques", June 2013
- IEC 62209-1, "Procedure to measure the Specific Absorption Rate (SAR) for hand-held devices used in close proximity to the ear (frequency range of 300 MHz to 3 GHz)", February 2005
- KDB 865664, "SAR Measurement Requirements for 100 MHz to 6 GHz"

**Additional Documentation:**

- DASY4/5 System Handbook

**Methods Applied and Interpretation of Parameters:**

- Measurement Conditions:** Further details are available from the Validation Report at the end of the certificate. All figures stated in the certificate are valid at the frequency indicated.
- Antenna Parameters with TSL:** The dipole is mounted with the spacer to position its feed point exactly below the center marking of the flat phantom section, with the arms oriented parallel to the body axis.
- Feed Point Impedance and Return Loss:** These parameters are measured with the dipole positioned under the liquid filled phantom. The impedance stated is transformed from the measurement at the SMA connector to the feed point. The Return Loss ensures low reflected power. No uncertainty required.
- Electrical Delay:** One-way delay between the SMA connector and the antenna feed point. No uncertainty required.
- SAR measured:** SAR measured at the stated antenna input power.
- SAR normalized:** SAR as measured, normalized to an input power of 1 W at the antenna connector.
- SAR for nominal TSL parameters:** The measured TSL parameters are used to calculate the nominal SAR result.

The reported uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor  $k=2$ , which for a normal distribution corresponds to a coverage probability of approximately 95%.

|                                |                                                                 |                                            |            |
|--------------------------------|-----------------------------------------------------------------|--------------------------------------------|------------|
| <b>SONY</b>                    | Sony Mobile Communications (China) Co.,<br>Ltd. Test Laboratory | Report No.: TARC-PM-0821-BV-<br>SAR-FCC-02 |            |
| PM-0821-BV SAR FCC Test Report |                                                                 | Edition 2                                  | Revision 0 |

### Measurement Conditions

DASY system configuration, as far as not given on page 1.

|                              |                        |             |
|------------------------------|------------------------|-------------|
| DASY Version                 | DASY5                  | V52.8.7     |
| Extrapolation                | Advanced Extrapolation |             |
| Phantom                      | Modular Flat Phantom   |             |
| Distance Dipole Center - TSL | 15 mm                  | with Spacer |
| Zoom Scan Resolution         | dx, dy, dz = 5 mm      |             |
| Frequency                    | 835 MHz $\pm$ 1 MHz    |             |

### Head TSL parameters

The following parameters and calculations were applied.

|                                         | Temperature         | Permittivity   | Conductivity         |
|-----------------------------------------|---------------------|----------------|----------------------|
| Nominal Head TSL parameters             | 22.0 °C             | 41.5           | 0.90 mho/m           |
| Measured Head TSL parameters            | (22.0 $\pm$ 0.2) °C | 40.5 $\pm$ 6 % | 0.93 mho/m $\pm$ 6 % |
| Head TSL temperature change during test | < 0.5 °C            | ----           | ----                 |

### SAR result with Head TSL

|                                                         |                    |                              |
|---------------------------------------------------------|--------------------|------------------------------|
| SAR averaged over 1 cm <sup>3</sup> (1 g) of Head TSL   | Condition          |                              |
| SAR measured                                            | 250 mW input power | 2.43 W/kg                    |
| SAR for nominal Head TSL parameters                     | normalized to 1W   | 9.43 W/kg $\pm$ 17.0 % (k=2) |
| SAR averaged over 10 cm <sup>3</sup> (10 g) of Head TSL | condition          |                              |
| SAR measured                                            | 250 mW input power | 1.57 W/kg                    |
| SAR for nominal Head TSL parameters                     | normalized to 1W   | 6.14 W/kg $\pm$ 16.5 % (k=2) |

### Body TSL parameters

The following parameters and calculations were applied.

|                                         | Temperature         | Permittivity   | Conductivity         |
|-----------------------------------------|---------------------|----------------|----------------------|
| Nominal Body TSL parameters             | 22.0 °C             | 55.2           | 0.97 mho/m           |
| Measured Body TSL parameters            | (22.0 $\pm$ 0.2) °C | 54.0 $\pm$ 6 % | 1.00 mho/m $\pm$ 6 % |
| Body TSL temperature change during test | < 0.5 °C            | ----           | ----                 |

### SAR result with Body TSL

|                                                         |                    |                              |
|---------------------------------------------------------|--------------------|------------------------------|
| SAR averaged over 1 cm <sup>3</sup> (1 g) of Body TSL   | Condition          |                              |
| SAR measured                                            | 250 mW input power | 2.39 W/kg                    |
| SAR for nominal Body TSL parameters                     | normalized to 1W   | 9.30 W/kg $\pm$ 17.0 % (k=2) |
| SAR averaged over 10 cm <sup>3</sup> (10 g) of Body TSL | condition          |                              |
| SAR measured                                            | 250 mW input power | 1.56 W/kg                    |
| SAR for nominal Body TSL parameters                     | normalized to 1W   | 6.11 W/kg $\pm$ 16.5 % (k=2) |

|                                |                                                                 |                                        |            |
|--------------------------------|-----------------------------------------------------------------|----------------------------------------|------------|
| <b>SONY</b>                    | Sony Mobile Communications (China) Co.,<br>Ltd. Test Laboratory | Report No.: TARC-PM-0821-BV-SAR-FCC-02 |            |
| PM-0821-BV SAR FCC Test Report |                                                                 | Edition 2                              | Revision 0 |

## Appendix

### Antenna Parameters with Head TSL

|                                      |                                |
|--------------------------------------|--------------------------------|
| Impedance, transformed to feed point | 52.0 $\Omega$ - 2.0 j $\Omega$ |
| Return Loss                          | - 31.1 dB                      |

### Antenna Parameters with Body TSL

|                                      |                                |
|--------------------------------------|--------------------------------|
| Impedance, transformed to feed point | 46.5 $\Omega$ - 4.5 j $\Omega$ |
| Return Loss                          | - 24.6 dB                      |

### General Antenna Parameters and Design

|                                  |          |
|----------------------------------|----------|
| Electrical Delay (one direction) | 1.388 ns |
|----------------------------------|----------|

After long term use with 100W radiated power, only a slight warming of the dipole near the feedpoint can be measured.

The dipole is made of standard semirigid coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC-signals. On some of the dipoles, small end caps are added to the dipole arms in order to improve matching when loaded according to the position as explained in the "Measurement Conditions" paragraph. The SAR data are not affected by this change. The overall dipole length is still according to the Standard.

No excessive force must be applied to the dipole arms, because they might bend or the soldered connections near the feedpoint may be damaged.

### Additional EUT Data

|                 |                   |
|-----------------|-------------------|
| Manufactured by | SPEAG             |
| Manufactured on | November 27, 2006 |

|                                |                                                                 |                                        |            |
|--------------------------------|-----------------------------------------------------------------|----------------------------------------|------------|
| <b>SONY</b>                    | Sony Mobile Communications (China) Co.,<br>Ltd. Test Laboratory | Report No.: TARC-PM-0821-BV-SAR-FCC-02 |            |
| PM-0821-BV SAR FCC Test Report |                                                                 | Edition 2                              | Revision 0 |

## DASY5 Validation Report for Head TSL

Date: 21.02.2014

Test Laboratory: SPEAG, Zurich, Switzerland

**DUT: Dipole 835 MHz; Type: D835V2; Serial: D835V2 - SN: 4d061**

Communication System: UID 0 - CW; Frequency: 835 MHz

Medium parameters used:  $f = 835 \text{ MHz}$ ;  $\sigma = 0.93 \text{ S/m}$ ;  $\epsilon_r = 40.5$ ;  $\rho = 1000 \text{ kg/m}^3$

Phantom section: Flat Section

Measurement Standard: DASY5 (IEEE/IEC/ANSI C63.19-2007)

DASY52 Configuration:

- Probe: ES3DV3 - SN3205; ConvF(6.22, 6.22, 6.22); Calibrated: 30.12.2013;
- Sensor-Surface: 3mm (Mechanical Surface Detection)
- Electronics: DAE4 Sn601; Calibrated: 25.04.2013
- Phantom: Flat Phantom 4.9L; Type: QD000P49AA; Serial: 1001
- DASY52 52.8.7(1137); SEMCAD X 14.6.10(7164)

### Dipole Calibration for Head Tissue/Pin=250 mW, d=15mm/Zoom Scan (7x7x7)/Cube 0:


Measurement grid: dx=5mm, dy=5mm, dz=5mm

Reference Value = 58.079 V/m; Power Drift = -0.07 dB

Peak SAR (extrapolated) = 3.66 W/kg


**SAR(1 g) = 2.43 W/kg; SAR(10 g) = 1.57 W/kg**

Maximum value of SAR (measured) = 2.85 W/kg



|                                                                                   |                                                                 |                                        |            |
|-----------------------------------------------------------------------------------|-----------------------------------------------------------------|----------------------------------------|------------|
|  | Sony Mobile Communications (China) Co.,<br>Ltd. Test Laboratory | Report No.: TARC-PM-0821-BV-SAR-FCC-02 |            |
| PM-0821-BV SAR FCC Test Report                                                    |                                                                 | Edition 2                              | Revision 0 |

Impedance Measurement Plot for Head TSL



|                                |                                                                 |                                        |            |
|--------------------------------|-----------------------------------------------------------------|----------------------------------------|------------|
| <b>SONY</b>                    | Sony Mobile Communications (China) Co.,<br>Ltd. Test Laboratory | Report No.: TARC-PM-0821-BV-SAR-FCC-02 |            |
| PM-0821-BV SAR FCC Test Report |                                                                 | Edition 2                              | Revision 0 |

## DASY5 Validation Report for Body TSL

Date: 20.02.2014

Test Laboratory: SPEAG, Zurich, Switzerland

**DUT: Dipole 835 MHz; Type: D835V2; Serial: D835V2 - SN: 4d061**

Communication System: UID 0 - CW; Frequency: 835 MHz

Medium parameters used:  $f = 835 \text{ MHz}$ ;  $\sigma = 1 \text{ S/m}$ ;  $\epsilon_r = 54$ ;  $\rho = 1000 \text{ kg/m}^3$

Phantom section: Flat Section

Measurement Standard: DASY5 (IEEE/IEC/ANSI C63.19-2007)

DASY52 Configuration:

- Probe: ES3DV3 - SN3205; ConvF(6.09, 6.09, 6.09); Calibrated: 30.12.2013;
- Sensor-Surface: 3mm (Mechanical Surface Detection)
- Electronics: DAE4 Sn601; Calibrated: 25.04.2013
- Phantom: Flat Phantom 4.9L; Type: QD000P49AA; Serial: 1001
- DASY52 52.8.7(1137); SEMCAD X 14.6.10(7164)

### Dipole Calibration for Body Tissue/Pin=250 mW, d=15mm/Zoom Scan (7x7x7)/Cube 0:


Measurement grid:  $dx=5\text{mm}$ ,  $dy=5\text{mm}$ ,  $dz=5\text{mm}$

Reference Value = 54.667 V/m; Power Drift = 0.02 dB

Peak SAR (extrapolated) = 3.56 W/kg


**SAR(1 g) = 2.39 W/kg; SAR(10 g) = 1.56 W/kg**

Maximum value of SAR (measured) = 2.80 W/kg



|                                                                                   |                                                                 |                                        |            |
|-----------------------------------------------------------------------------------|-----------------------------------------------------------------|----------------------------------------|------------|
|  | Sony Mobile Communications (China) Co.,<br>Ltd. Test Laboratory | Report No.: TARC-PM-0821-BV-SAR-FCC-02 |            |
| PM-0821-BV SAR FCC Test Report                                                    |                                                                 | Edition 2                              | Revision 0 |

Impedance Measurement Plot for Body TSL





|                                                                                   |                                                                 |                                        |            |
|-----------------------------------------------------------------------------------|-----------------------------------------------------------------|----------------------------------------|------------|
|  | Sony Mobile Communications (China) Co.,<br>Ltd. Test Laboratory | Report No.: TARC-PM-0821-BV-SAR-FCC-02 |            |
| PM-0821-BV SAR FCC Test Report                                                    |                                                                 | Edition 2                              | Revision 0 |

## APPENDIX F: GLOSSARY

|          |   |                                                |
|----------|---|------------------------------------------------|
| BW       | - | Bandwidth                                      |
| DTS      | - | Distributed Transmission System                |
| DUT      | - | Device under Test                              |
| EUT      | - | Equipment under Test                           |
| FCC      | - | Federal Communication Commission               |
| FCC ID   | - | Company Identifier at FCC                      |
| HW       | - | Hardware                                       |
| IC       | - | Industry Canada                                |
| Inv. No. | - | Inventory number                               |
| LTE      | - | Long Term Evolution                            |
| N/A      | - | Not Applicable                                 |
| NP       | - | Not Performed                                  |
| PCE      | - | Personal Consumption Expenditure               |
| OET      | - | Office of Engineering and Technology           |
| RB       | - | Resource Block(s)                              |
| SAR      | - | Specific Absorption Rate                       |
| S/N      | - | Serial Number                                  |
| SPLSR    | - | SAR-to-(peak-locations spacing) ratio          |
| SW       | - | Software                                       |
| UNII     | - | Unlicensed National Information Infrastructure |

|                                                                             |                                            |            |
|-----------------------------------------------------------------------------|--------------------------------------------|------------|
| <b>SONY</b> Sony Mobile Communications (China) Co.,<br>Ltd. Test Laboratory | Report No.: TARC-PM-0821-BV-<br>SAR-FCC-02 |            |
| PM-0821-BV SAR FCC Test Report                                              | Edition 2                                  | Revision 0 |

This page has been left intentionally blank.