



**Laerdal**  
helping save lives

## Permissive Change Request Letter

Date: November 24, 2023

Federal Communications Commission  
7435 Oakland Mills Road  
Columbia, MD 21046

FCC ID: QHQ-LB1YM  
Original Grant Date: October 30, 2023

To Whom It May Concern:

We, the undersigned, hereby attest to the fact that we will apply the class II change.

The major change filed under this application is:

1. Additional platform added (Product: SimPad PLUS 2, Brand: Laerdal Medical AS, Model: 204-00150).
2. The device in original application was certified with below antennas:

Ant.	Supplier	Model No.	Type	Antenna Gain (dBi)	
				BT & WiFi 2.4GHz	WiFi 5GHz
0	Molex	146153	Dipole	3.2	4.25
1	Molex	146187	Dipole	3.4	4.75

This class II change is to add Dipole antennas with lower gain as the following table.

Ant.	Brand Name	Model No.	Type	Antenna Gain (dBi)		
				BT	WiFi 2.4GHz	WiFi 5GHz
0	SANAV	GEPH-062-1	Dipole	-	-1.99	U-NII 1: -1.33 U-NII 2A: -1.33 U-NII 2C: -5.36 U-NII 3: -3.36
1	SANAV	GEPH-056	Dipole	3.04	3.04	U-NII 1: -2.62 U-NII 2A: -2.62 U-NII 2C: -2.21 U-NII 3: -1.93

Should you have any enquires or comments, please have my best attention.

Sincerely,

DocuSigned by:

*Mari Kaada*

Signature

50D90013D6AA495

Contact Name: Mari Kaada

Job Title: Corporate Director Q&R

Email: mari.kaada@laerdal.com