

BIOTRONIK GmbH & Co. KG • Postfach 47 02 55 • 12311 Berlin

Federal Communications Commission
Equipment Authorization Branch
7435 Oakland Mills Road
Columbia, MD 21046
USA

Berlin, February 1, 2007

Re: BIOTRONIK Patient Device CardioMessenger, FCC ID: QRICM06-1
granted July 27, 2006

Dear Madam or Sir

This Class II permissive change request applies to FCC ID: QRICM06-1.
The antenna matching network has been changed to allow a low VSWR connection to 50 Ohms cable for manufacturing.
The bottom antenna has been changed.
This change will be incorporated in all units once accepted by the Commission.
In support of the Class II change, test reports for radiated emissions as well as the new antenna specification and circuit layout including the parts list are being submitted with the application.

Sincerely



Gunnar Boersch
Team Manager Regulatory Affairs