



✓ *Where People Make the Difference*

---

March 5, 2008

AmericanTCB  
6731 Whittier Avenue  
McLean VA 22101

Attn: Director of Certification

Re: Compliance Statement in Device Manual (Instructions For Use) for Nonin Model "Saber"  
FCC ID: RG9SABER / IC: 4743A-SABER

The model "Saber" Bluetooth Communication Module is installed in our company's own Medical Electrical (ME) Equipment as original equipment. We do not supply a manual or advertising literature with the Saber module but we do notify our Documentation Department of the required statements that must be placed in the ME Equipment Instructions For Use (IFU), meeting the requirements of the Federal Communications Commission CFR 47, Part 15, Paragraph 15.19(a)(3) and Industry Canada RSS-GEN, Paragraph 7.1.5:

A representative copy of this notification memo is attached.

Sincerely,

A handwritten signature in cursive script, appearing to read "Harold Rudnick".

Harold Rudnick, NCE  
Sr. Quality Engineer  
TEL 763-553-9968  
FAX 763-553-7807  
harold.rudnick@nonin.com



✓ *Where People Make the Difference*

# Memo

**To:** John Gawron, Documentation Manager  
**From:** Harold Rudnick, Sr. Quality Engineer  
**Date:** March 5, 2008  
**RE:** Radio equipment compliance statements

The model "Saber" Bluetooth Communication Module is installed in our ME Equipment as original equipment. We do not supply a manual with the Saber module but we request that the following required statements be placed in the ME Equipment manual to meet the requirements of the Federal Communications Commission CFR 47, Part 15, and Industry Canada RSS-GEN.

#### **COMPLIANCE STATEMENT**

"This device complies with Part 15 of the FCC rules. Operation is subject to the following two conditions: 1) This device may not cause harmful interference, and, 2) This device must accept any interference received including interference that may cause undesired operation."

#### **MODIFICATION CAUTION STATEMENT:**

"CAUTION TO USER: Changes or modifications not expressly approved by the party responsible for compliance could void the user's authority to operate the equipment."

Sincerely

A handwritten signature in cursive script that reads "Harold Rudnick".

Harold Rudnick  
Sr. Quality Engineer