

HAC_E_Dipole_835_130630

DUT: HAC Dipole 835 MHz

Communication System: CW; Frequency: 835 MHz; Duty Cycle: 1:1

Medium: Air Medium parameters used: $\sigma = 0 \text{ S/m}$, $\epsilon_r = 1$; $\rho = 0 \text{ kg/m}^3$

Ambient Temperature : 23.2 °C

DASY5 Configuration:

- Probe: ER3DV6 - SN2476; ConvF(1, 1, 1); Calibrated: 2012-12-12;
- Sensor-Surface: (Fix Surface)
- Electronics: DAE4 Sn679; Calibrated: 2013-1-16
- Phantom: HAC Test Arch with AMCC; Type: SD HAC P01 BA;
- Measurement SW: DASY52, Version 52.8 (5); SEMCAD X Version 14.6.8 (7028)

E Scan - measurement distance from the probe sensor center to CD835 = 15mm/Hearing Aid Compatibility Test at 15mm distance (41x361x1): Interpolated grid:

dx=0.5000 mm, dy=0.5000 mm

Device Reference Point: 0, 0, -6.3 mm

Reference Value = 103.7 V/m; Power Drift = -0.01 dB

PMR not calibrated. PMF = 1.000 is applied.

E-field emissions = 106.7 V/m

Average value of Total=(106.7+106.0)/2=106.35 V/m

PMF scaled E-field



Grid 1 M4 105.3 V/m	Grid 2 M4 106.7 V/m	Grid 3 M4 104.9 V/m
Grid 4 M4 63.11 V/m	Grid 5 M4 63.81 V/m	Grid 6 M4 62.47 V/m
Grid 7 M4 104.7 V/m	Grid 8 M4 106.0 V/m	Grid 9 M4 104.1 V/m

Cursor:

Total = 106.7 V/m

E Category: M4

Location: 0.5, -70, 9.7 mm



0 dB = 106.7 V/m = 40.56 dBV/m

HAC_E_Dipole_1880_130630

DUT: HAC Dipole 1880 MHz

Communication System: CW; Frequency: 1880 MHz; Duty Cycle: 1:1

Medium: Air Medium parameters used: $\sigma = 0 \text{ S/m}$, $\epsilon_r = 1$; $\rho = 0 \text{ kg/m}^3$

Ambient Temperature : 23.2 °C

DASY5 Configuration:

- Probe: ER3DV6 - SN2476; ConvF(1, 1, 1); Calibrated: 2012-12-12;
- Sensor-Surface: (Fix Surface)
- Electronics: DAE4 Sn679; Calibrated: 2013-1-16
- Phantom: HAC Test Arch with AMCC; Type: SD HAC P01 BA;
- Measurement SW: DASY52, Version 52.8 (4); SEMCAD X Version 14.6.8 (7028)

E Scan - measurement distance from the probe sensor center to CD1880 = 15mm/Hearing Aid Compatibility Test at 15mm distance (41x181x1): Interpolated grid:

dx=0.5000 mm, dy=0.5000 mm

Device Reference Point: 0, 0, -6.3 mm

Reference Value = 169.3 V/m; Power Drift = 0.03 dB

PMR not calibrated. PMF = 1.000 is applied.

E-field emissions = 94.17 V/m

Average value of Total=(88.92+94.17)/2=91.545 V/m

PMF scaled E-field



Grid 1 M3 87.14 V/m	Grid 2 M3 88.92 V/m	Grid 3 M3 87.82 V/m
Grid 4 M3 72.36 V/m	Grid 5 M3 73.09 V/m	Grid 6 M3 71.98 V/m
Grid 7 M3 92.44 V/m	Grid 8 M3 94.17 V/m	Grid 9 M3 92.49 V/m

Cursor:

Total = 94.17 V/m

E Category: M3

Location: 0, 38, 9.7 mm



0 dB = 94.17 V/m = 39.48 dBV/m