



775 Montague Expressway
Milpitas, CA 95035
Tel: 408-526-1188
Fax: 408-526-1088
Email: sales.eaw@us.bureauveritas.com

FCC CLASS II PERMISSIVE CHANGE REQUEST LETTER

Reason for Amendment (current / obsolete)	Revision History		Approved Date
	From	To	
Initial Release (obsolete)	1.0	1.0	Feb-20-2012
Update company logo (current)	1.0	2.0	Jul-29-2019

FCC Class II Permissive Change Request Letter

Date: 03/04/2022

To FCC:

RE: FCC Permissive II Change Request for Company: ICU Medical, Inc. FCC ID: STJ-SDMAC

We are submitting an application for a class II permissive change to the FCC approval of the Company name: ICU Medical, Inc., product description: SDIO Wireless Module (FCC: **STJ-SDMAC**, Original Grant Date: 8/25/2021). The transmitter module itself has not changed. Here are the changes:


Additional PCB antenna (Peak Gain: 2.4GHz 3.00 dB, 5GHz 4.45 dB)

The device is operated in the portable condition

Bluetooth is disabled with product firmware

Mounted to ICU Medical Infusion Pump HMN: Plum Duo and operated throughout the test

Sincerely,



Title Director of BSP and Cybersecurity

Name: Chaitanya Srinivasamurthy

Address 600 N Field Dr., Lake Forrest, IL 60045

Signature