



LSRESEARCH, LLC

Wireless Product Development

Date: December 28, 2012

Re: FCC ID: THL-IMDAMSG3

Applicant: Angel Medical Systems, Inc.

Correspondence Reference Number: 43103

Form 731 Confirmation Number: EA215127

Response Letter

To Diane Poole:

Thank you for the review and comments. Based on the referenced KDB publication (447498) section 4.2.4 it appears that SAR test exclusion *may* be applied. We believe that the submitted SAR analysis exceeds* the requirements of the optional exclusion and should be accepted as proof of SAR compliance.

*The reported SAR analysis levels are significantly lower than the SAR limits as well as the submitted test report indicating the maximum power when measured in a radiated fashion much lower than 1.0 mW.

Sincerely,

A handwritten signature in black ink that reads "Thomas T. Smith". The signature is written in a cursive, flowing style.

Thomas T. Smith
Manager of EMC Test Services