



To:  
Federal Communications Commission  
Equipment Authorization Division  
7435 Oakland Mills Road  
Columbia, MD 21048  
U.S.A.

**Widex A/S**  
Nymoellevej 6  
DK-3540 Lyngé  
Denmark  
Tel.: (+45) 44 35 56 00  
Fax: (+45) 44 35 56 01  
widex@widex.com  
<http://www.widex.com>  
CVR. Nr. 1577 1100

Date: 29 October 2019

### **FCC Labeling of Widex model MRB2D**

FCC ID: TTY-MRB2D

#### To Whom It May Concern:

The dimensions of the hearing aid are too small to allow the FCC ID to be printed on the device in a font size that is readable. Consequently, the FCC ID is printed in the user instructions together with the FCC regulatory statements.

Justification is provided by 3 photographs below that show the MRB2D in 3 different angles together with a printout of the FCC ID in font size 4 point that is the minimum readable font size as defined in FCC OET Labelling Guidelines.



Thank you for your attention to this matter.

Sincerely,  
*Widex A/S*

Mads Kronborg Liin  
RF Design Engineer, Wireless system integration