

APPENDIX A MEASUREMENT INSTRUMENTATION DETAILS

EQUIPMENT TYPE	MAKE/MODEL SERIAL NUMBER	LAST CAL. DD/MM/YY	DUE DATE DD/MM/YY	CAL. INTERVAL
EMI RECEIVER	HP 8546A, Sn: 3549A00290	04/04/06	04/04/07	1 YEAR *2
ANTENNAS	EMCO 93110B BICONICAL 20 - 300 MHz Sn. 9804-3092	24/06/06	24/06/07	1 YEAR *1
	EMCO 93146A LOG PERIODIC 200 -1000MHz Sn. 5033	15/06/06	15/06/07	1 YEAR *1
	EMCO LOOP ANTENNA 9 kHz – 30 MHz Sn: 9108-2660	29/04/05	29/04/08	3 YEAR *1

Note *1. In-house calibration. Refer to Quality Manual.

Note *2. NATA calibration by Agilent Technologies (Aust) Pty Ltd