

To : Underwriters Laboratory Inc.

330 Pfingsten Road

Northbrook, IL 60062-2096 U.S.A

Attn : TCB Certifier

Date : July 4, 2007

Subject : Application of Class II Permissive change Authorization FCC ID: VAXINTROMEDIC

We, IntroMedic Co.,Ltd. hereby apply for FCC Class II Permissive change authorization for our product Capsule Endoscope System FCC ID: VAXINTROMEDIC.

The following change have been made to the original certified product.

Change Isolation Method.

- Delete DCDC Converter
- Connect DCDC Converter Input to Output
- Insert Safety Capacitor for AP Isolation

The product is identical to the previously certified Capsule Endoscope System except for the changes described above.

All other mechanical and electrical ratings are the same as original product.

Attached are the Test report, Internal/External photos, block diagram, ID Label info and User manual.

If you have any question on this application, please contact the undersigned.

Sincerely,



(Signature)

Typed name : Andrew Shim / President

IntroMedic Co.,Ltd.

Suite 1104, E&C Venture Dream Tower 6-Cha

197-28 Guro-Dong, Guro-Gu, Seoul, KOREA 152-719

www.intromedic.com

Phone : +82-2-801-9300

Fax : +82-2-801-9330