

Federal Communications Commission
7435 Oakland Mills Road
Columbia, MD 21046

Date: 23 July 2024

Subject: Label location for: FCC ID: VNP-ME1700


Ladies and gentlemen:

Pursuant to section §2.925 of 47 C.F.R. the MED-EL audio processor for a cochlear implant system, which is subject of this filing, shall bear a permanently affixed, readily visible label listing the information as specified in §2.926 and §15.19(a) of 47 C.F.R.

This audio processor is part of an active implantable medical device system and is worn behind the ear of a user.

Pursuant to sections §2.925(f) and §15.19(a)(5) this device is too small to carry a legible FCC ID and compliance statement on the device itself. The FCC ID and compliance statement will be included in the user manual. The FCC ID will also be placed on the device packaging.

Sincerely,



(Bernhard Stoebich, Team Leader Electronic Design, R&D)

MED-EL Elektromedizinische Geräte Gesellschaft m.b.H.

Fürstenweg 77a
6020 Innsbruck
Austria
Tel +43 5 77 88
Fax +43 512 29 33 81
office@medel.com

FN 48608h, DVR 1001795
Register Court:
Landesgericht Innsbruck
UID-No. ATU31722507
www.medel.com